

# CERTIFICATE OF CALIBRATION



**CHOKSI**  
LABORATORIES LIMITED

CALIBRATION CERTIFICATE NUMBER:03900003-07/21-22

Issued from : Choksi Laboratories Limited  
Survey No. 9/1, Balaji Tulsayana Industrial Estate,  
Kumedi, Off. M.R. 10 Toll Naka,  
Indore - 452010 (M.P.), INDIA

<b>EQUIPMENT NAME :</b> Fully Automatic Biochemistry Analyzer(Temperature)	<b>CALIBRATION DUE DATE RECOMMENDED</b>	<b>PAGE NO.</b>
<b>RECEIVED ON</b> 28.05.2021	<b>CALIBRATED ON</b> 31.05.2021	1 of 1

**I. Calibrated for :**  
PRESTIGE MEDICAL INVESTIGATION, INDORE  
12/2, NEW AGRAWAL NAGAR, AGRASEN SQUARE  
Indore, Madhya Pradesh  
India - 452001

**II. Description and Identification of Instrument :**

<b>Sample Id:</b>	03900003-07
<b>Make:</b>	ABM
<b>Model:</b>	ABM-200
<b>Type:</b>	SUN6020 1655T
<b>Serial No.</b>	Not mentioned

<b>Range:</b>	30.0 to 45.0 °C
<b>Calibrated Range:</b>	30.1 to 45.1 °C
<b>Least Count:</b>	0.1 °C
<b>Condition on receipt:</b>	Satisfactory
<b>Identification No. :</b>	Not mentioned

<b>Temperature :</b>	25 ± 5 °C
<b>Relative Humidity :</b>	45 to 75 % RH

**III. Environment Condition-**

**IV. Standard/Calibrator and Measuring Equipment Used :**

Nomenclature	Range	Uncertainty	Certificate No.	Traceability	Validity
Digital thermometer with sensor	-79.8 to 299.6 °C	± 0.142 to 0.350 °C	M-200629-21-1	CC-2634	16.07.2021

**V. Results: (Temperature)**

Thermal Calibration					
Serial No.	Unit	UUC*Average	Standard Average	Correction	Uncertainty (For k=2, 95.45 % C.L. ) : ±
1	°C	30.06	30.44	0.38	0.82
2	°C	35.08	35.02	-0.06	0.82
3	°C	37.04	37.06	0.02	0.82
4	°C	40.06	40.70	0.64	0.82
5	°C	45.08	45.10	0.02	0.82

**Note :** 01.UUC\* : Unit Under Calibration  
: 02. Certificate shall not be reproduced in full without the written approval of the laboratory  
: 03. The results relate only to the item calibrated.

**Other Details**

01.Certificate issue date : 02/06/2021  
02.Location of calibration: Site  
03. The results relate only to the item calibrated

**VI. Inference:**

01. The correction is positive except when indicated other wise It has to be algebraically added to the UUC\* reading in order to get the actual reading .

**VII. Reference:**

01. The Calibration is carried out as per the CLL Calibration Procedure  
No. CLL-IDR-SWP-CAL-00041

02. All the values have been rounded off as per IS:2-1960.

03.The laboratory is NABL Accredited having certificate No. CC-2089, ULR NO.,CC208921000001237F

**VIII. Details of Adjustment Made (If any)**

Not Done

**CALIBRATED BY** IDR00910

**AUTHORISED SIGNATORY**

Signature Not Verified  
Mitesh J. Doshi IDR01151  
02.06.2021 14:24

\*\*END OF CALIBRATION CERTIFICATE\*\*