

## VIEW LAB MONTHLY SUMMARY

Lab Name	PRESTIGE MEDICAL INVESTIGATIONS		Details About Robust Analysis								
Lab No	5672		Detail About Monthly Summary								
Month	October		Detail about SDI								
Year	2021										
Constituent Group	Chemistry I										
Click on the analyte to view Graphical Data			<a href="#">All Analyser Result</a>		<a href="#">Print</a>						
			Date of Result Entered : 16/10/2021								
			Date of Report Published : 01/11/2021								
Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U	
						CV	SD				
1	GLUCOSE	GOD-POD	Any Analyser (Automation / Semi Automation )	702	315.96	8.56	27.03	360 mg/dl	1.63	2.04	
2	UREA	Others ( Urease Berthelot / Ned Dye )	Any Analyser (Automation / Semi Automation )	196	44.22	11.73	5.19	42 mg/dl	-0.43	0.74	
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Any Analyser (Automation / Semi Automation )	486	1.13	13.39	0.15	1.13 mg/dl	0.00	0.01	
4	T.BILIRUBIN	Diazonium salt ( Colorimetric ) / Jendrassik	Any Analyser (Automation / Semi Automation )	557	6.61	11.91	0.79	6.3 mg/dl	-0.39	0.07	
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation )	677	5.29	9.77	0.52	5.92 g/dl	1.22	0.04	
6	ALBUMIN	BCG - colorimetric	Any Analyser (Automation / Semi Automation )	657	3.27	8.41	0.28	3.75 g/dl	1.75	0.02	
7	CALCIUM	Arsenazo III	Any Analyser (Automation / Semi Automation )	506	9.53	8.10	0.77	9.8 mg/dl	0.35	0.07	
8	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Any Analyser (Automation / Semi Automation )	319	5.40	11.35	0.61	5.01 mg/dl	-0.64	0.07	
9	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation )	681	5.40	16.16	0.87	5.9 mg/dl	0.57	0.07	
10	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation )	677	105.70	11.01	11.63	115 mg/dl	0.80	0.89	
11	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation )	597	97.58	14.69	14.34	74 mg/dl	-1.64	1.17	
12	HDL CHO	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation )	438	27.23	16.03	4.37	29 mg/dl	0.41	0.42	
13	SODIUM	ISE - Indirect	Roche C111, C311, Integra 400 plus, C501, C502, C702	303	135.40	2.97	4.02	138 mmol/L	0.65	0.46	
14	POTASSIUM	ISE - Indirect	Roche C111, C311, Integra 400 plus, C501, C502, C702	306	3.28	4.84	0.16	3.92 mmol/L	4.03	0.02	
15	CHLORIDE	ISE - Indirect	Roche C111, C311, Integra 400 plus, C501, C502, C702	233	107.39	3.93	4.22	108 mmol/L	0.14	0.55	
16	AST	UV kinetic with PLP ( P-	Any Analyser	625	206.93	13.92	28.82	172 U/L	-1.21	2.31	

		5-P )	(Automation / Semi Automation )							
17	ALT	UV kinetic with PLP ( P- 5-P )	Any Analyser (Automation / Semi Automation )	641	75.62	20.05	15.16	58 U/L	-1.16	1.20
18	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation )	426	93.76	18.94	17.76	135 U/L	2.32	1.72
19	AMYLASE	CNPG3	Any Analyser (Automation / Semi Automation )	407	125.52	17.70	22.21	135 U/L	0.43	2.20
20	MAGNESIUM	Xylidyl Blue	Any Analyser (Automation / Semi Automation )	157	1.84	17.49	0.32	2.39 mg/dl	1.71	0.05

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

### Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out