



CENTRE for HIV TESTING
State Reference Laboratory for HIV
Department of Microbiology
Lady Hardinge Medical College & Associated Hospitals
New Delhi

Ref. No. *MIC/2021/904/CHT-LHMC/167*

Date: 02/11/21

EQAS REPORT (SRL to ICTCs/PPTCTs)

Round: I

Name of Proficiency Testing provider: SRL, CHT-LHMC Date of panel distribution: 28/10/21

Name of the participating laboratory: Dr. RML Hospital (ICTC)

Date of result received: 02/11/21

Sample code	Standard result	Results obtained by participating laboratory	% matching
A4	Positive for HIV-1 ab	Reactive for HIV-1 abs	100
A1	Negative for HIV ab	Non-reactive for HIV abs	100
A2	Negative for HIV ab	Non-reactive for HIV abs	100
A3	Positive for HIV-1 ab	Reactive for HIV-1 abs	100

Remarks: Performance is satisfactory.

Dr. V-S-Randhawa
Name of Laboratory In-Charge:


Signature of Laboratory In-Charge:

*** End of Report***



CENTRE for HIV TESTING
State Reference Laboratory for HIV
Department of Microbiology
Lady Hardinge Medical College & Associated Hospitals
New Delhi

Ref. No. MIC/2021/299/CHT-LHMC/110

Date: 27/03/21

EQAS REPORT (SRL to ICTCs/PPTCTs)

Round: II

Name of Proficiency Testing provider: SRL, CHT-LHMC Date of panel distribution: 20/03/21

Name of the participating laboratory: Dr. RML Hospital (ICTC)

Date of result received: 27/03/21

Sample code	Standard result	Results obtained by participating laboratory	% matching
A4	Negative for HIV ab	Negative for HIV antibody	100
A1	Positive for HIV-1 ab	Positive for HIV antibodies	100
A2	Negative for HIV ab	Negative for HIV antibody	100
A3	Positive for HIV-1 ab	Positive for HIV antibodies	100

Remarks: Performance is satisfactory.

Dr. V. S. Randhawa
Name of Laboratory In-Charge:

for *AMS*
Signature of Laboratory In-Charge:

*** End of Report***

mic/2021/299/CHT-LHMC/110
State Reference Laboratory (SRL)
Centre for HIV Testing (CHT)
Department of Microbiology
L.H.M.C., New Delhi-110004