



# UNIQUE CALIBRATION SOLUTIONS LLP

An ISO/IEC-17025:2017 / ACCREDITED BY NABL

UNIQUE  
CALIBRATION & SERVICES

Tel: 0484-4858563, 0484 - 7966984, 9971328563, 9061728563, 9778197603

E-mail: sales@uniquecalibration.com, website: www.uniquecalibration.com



CC-3102

CALIBRATION CERTIFICATE				
UNIQUE LAB REPORT NO (ULR NO) : CC310221000001889F				
SRF NO.	190	CERTIFICATE NO	UCSL/03-21/187_04	
User Name	M/s.CPC MEDICAL LABORATORY, SREEDEVI BUILDING,MAVOOR ROAD,BEHIND NANDILATH G-MART, ARAYIDATHUPALAM,KOZHICODE,KERALA-673004			
Device Under Calibration	CENTRIFUGE (RPM)		Equipment & Master Used For calibration	
Make / Model No.	ROTEK/M8E+AH1		01. DIGITAL TACHO METER (NON-CONTACT)	
Range	0 to 5500 rpm		Make : LUTRON	
Least Count	100 rpm		Model No. : DT -2234 C	
Instrument Id. No.	---		Certificate No. : BLR21/L/131-01	
Instrument Sr. No.	0868		Next Due Date : 08.03.2022	
Location	HEMATOLOGY LAB		Traceability : BLR	
Visual Inspection	OK			
Material Receipt Date	---			
Calibration Performed at	ON SITE			
Date of Calibration	25.03.2021			
Suggested Due Date of Calibration	25.03.2022			
Calibration Certificate Issue Date	29.03.2021			
ENVIRONMENTAL CONDITION				
Temperature	26.3°C			
Relative Humidity	55%RH			
Condition Of Equipment :	Un-Loaded			
OBSERVATION RESULTS				
SL. NO.	Nominal Value on UUC (rpm)	Observed Mean Value on MASTER (rpm)	% Error of Reading (±)	± Expanded Uncertainty At 95% Confidence Level (k=2)
1	1000.0	1001.6	0.16	0.5 % rdg
2	2000.0	2002.0	0.10	0.5 % rdg
3	3000.0	3001.8	0.06	0.5 % rdg
4	4000.0	4002.4	0.06	0.5 % rdg
5	5500.0	5503.0	0.05	0.5 % rdg
Calibration Procedure Based On	: UCSL/SOP/RPM			
NOTE :-				
1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.				
2. This report should not be reproduced except in full without our prior permission in writing.				
3. Calibration certificate without signature are not valid.				
4. All our certificates are Traceable to National standard.				
5. UUC : Unit under calibration , GUT : Gauge under testing.				
Calibrated By			Issued / Approved By (APPU K MANI) Technical Director	

Laboratory : 5/D3, EK Apartments, Model engineering college road Thrikkakara, Edappally, Cochin, Kerala-682021  
Registered Office : Plot No. 611-P, Sector - 38, Gurgaon Haryana - 122001

SALES \* SPARES \* SERVICE \* TESTING \* CALIBRATION \* MODERNIZATION