



Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area
Delhi - 110092

Phone No. : 09810943334 011-41764006, Email: multitech.technofab@gmail.com, mtedelhi2020@gmail.com,
Web. : www.multitechtechnofab.com



CC3225

CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/1604				Certificate Issued on: 02-08-2021							
Job / Identification No. : MTE/CAL/JOB/2466				Date of Calibration : 31-07-2021							
Item Receipt Date: 30-07-2021				Recommended Date of Next Calibration: 31-07-2022							
Calibration Certificate for : Micropipette						Item Location :					
CALIBRATED FOR		HARSHIT MEDICAL RESEARCH CENTRE Tiraha, near Varun Hotel, Wazidpur, , Jaunpur, Uttar Pradesh, 222002			Reference: Order thru Telephonic						
DISCIPLINE		Mechanical		GROUP		Volume					
Item's Description											
Name of the Item		Make / Model		Serial / I.D. No.		Working / Inst. Range		Least Count	Unit	Item Condition	
Micropipette		Erba / NA		NA / CHARAK/MP/2		10 to 100		1	µL	Working	
ENVIRONMENTAL CONDITION		Temperature (in °C):		25.5	± 2 °C	Humidity (in % RH):	51	± 10% RH	Atm. Pressure in (hPa) 992		
STANDARD REFERENCE DETAILS Traceable to National / International Standard											
Name of the Standard		Make / Model		Certificate No.		Calibration Agency / Traceability		Date of Calibration		Valid up to	
Semi Micro Balance		DANWER / DWA265		BBT/091/JAN/21		BLUEBOX TECHNO		07-01-2021		07-01-2022	
PRINCIPLE/METHODOLOGY OF CALIBRATION:				Lab Procedure No.: MTE-2/M/2.9			St. Procedure No.: ISO 8655-6				

RESULTS: In-Lab Calibration
All Readings are in µL, Specified Otherwise

U.U.C. READING	Volume @ 27°(ml) Chanel 1	UNCERTAINTY [^]
10	0.010061	± 0.06
50	0.050186	± 0.06
100	0.100349	± 0.06

[^]The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.

REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.

Note


- This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.
- The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor implied.
- Total liability of our organization is limited to the invoiced amount.
- Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
- In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

***** END OF REPORT *****

ULR : CC322521000001053F

Calibrate By
(Parveen Kumar) , parveen
Deputy Technical Manager

For and on behalf of
Multitech Technofab Enterprises


Laboratory Head