


LAB MONTHLY SUMMARY



Lab Name **YASH DIAGNOSTIC CENTRE** Lab No **14650**
 Month **October** Year **2021**
 Constituent Group **Chemistry I**



Date of Result Entered : 20/10/2021

Date of Report Published : 01/11/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Transasia / Erba	850	329.52	6.85	22.56	373.8 mg/dl	1.96	1.55
2	UREA	Urease UV / GLDH	Transasia / Erba	817	44.07	9.45	4.17	47.3 mg/dl	0.78	0.29
3	CREATININE	Enzymatic Colorimetric	Transasia / Erba	347	1.08	9.70	0.10	1.04 mg/dl	-0.38	0.01
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Transasia / Erba	787	6.68	9.07	0.61	6.82 mg/dl	0.23	0.04
5	T-PROTEIN	Biuret - Colorimetric	Transasia / Erba	797	5.49	7.47	0.41	5.44 g/dl	-0.12	0.03
6	ALBUMIN	BCG - colorimetric	Transasia / Erba	811	3.21	5.23	0.17	3.43 g/dl	1.31	0.01
7	CALCIUM	Arsenazo III	Transasia / Erba	576	9.76	8.19	0.80	9.8 mg/dl	0.05	0.07
8	URIC ACID	Enzymatic / Uricase Colorimetric	Transasia / Erba	807	5.66	9.43	0.53	5.9 mg/dl	0.45	0.04
9	CHOLESTEROL	CHOD-PAP	Transasia / Erba	838	100.17	9.14	9.15	116 mg/dl	1.73	0.63
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Transasia / Erba	736	79.80	12.65	10.09	74 mg/dl	-0.57	0.74
11	HDL CHO	Direct method / Enzymatic colorimetric	Transasia / Erba	624	26.59	12.09	3.22	24 mg/dl	-0.81	0.26
12	AST	UV kinetic with PLP (P-5-P)	Any Analyser (Automation / Semi Automation)	625	206.93	13.92	28.82	212.9 U/L	0.21	2.31
13	ALT	UV kinetic with PLP (P-5-P)	Any Analyser (Automation / Semi Automation)	641	75.62	20.05	15.16	74.2 U/L	-0.09	1.20
14	ALP	PNP AMP kinetic	Transasia / Erba	690	86.75	15.60	13.53	91 U/L	0.31	1.03

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Co-Ordinator
Christian Medical College, Vellore

******* End of Report *******