



## Calibration Certificate, Semi Automatic BioChemistry Analyzer

Name	STAR – 20 ( SEMIAUTO BIOCHEMISTRY ANALYZER)
Hospital	YASH DIAGNOSTIC CENTER , VISHESHARPUR PACHAHATIYA U.P. – (222001)
Date	20 July 2021

This is certify that STAR-20 (SEMIAUTO BIOCHEMISTRY ANALYZER) have calibrated & produced following data..

### Action Details:-

- AD VALUE - ( 49200 To 51550 ) Limits – 35000 to 55000
- OFFSET VALUE- ( 34 To 126 ) Limits- 0.0 to 4000
- Water Blank Abs - 0.005 Limits- < 0.01
- Temp Value - < 5% Variation Limits - < 5%

Instrument Identification				QC Passed	
Serial no.		Date of Calibration	20.07.2021	Yes	Yes
Model	STAR-20	Due Date of Calibration	19.07.2022	Yes	Yes

After calibration Q.C. Test for all level passed successfully.

Test results is coming ok	✓	Remarks	Done
---------------------------	---	---------	------

Service Engineer

