



# Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092  
Ph.: 011-42657237 Mob.: 9810804449, 8368671530  
E- mail: primac21@gmail.com



Format No.  
F01(7.8/01)

## CALIBRATION CERTIFICATE

ULR No. : CC233921000000662F	Calibration Done On : 23.01.2021	S.R.F.No. & Date : 0001/22 & 22.01.2021
Issue Date : 23.01.2021	Calibration Valid Upto : 22.01.2022	Page : 1 of 1
UUC Receipt Date : 22.01.2021	Calibrated for M/S.	
<p>Dr. Das Pathlab A-28, Lower Ground Floor, Ghanshyam Dass Jain Marg, South Extension Part-2, New Delhi-49</p>		

Identification on UUC (Unit Under Calibration)			
Instrument Name	Micro Pipette	Size/Range	10 - 100 µl
Least Count	1 µl	Condition of Item	Good
Manufactured By	Erba	Calibration Performed at	Lab
Manufacturer M/Sr.No.	YE175AB0104787	Location	Lab
Party ID Mark	--	Ref. Standard	ISO-8655-6

Calibration Procedure / Method : As per Calibration Procedure Document No. PCS/CP/01, Section No.CP-51

Environmental conditions (s) : Temperature: 23 ± 3 °C. Relative Humidity : 50 ± 10% RH

Standard & Major Equipment(s) Used :					
Standard Equipment	Range/ Serial No.	Calibrating Agency	Traceability Cert. No.	Certificate Date	Valid Up to
Standard Weight	0.001 g to 200 g	Weightronics	WMCL/E/2020-06/3369	03.06.2020	02.06.2023

Calibration Results at 27 °C :			
Nominal Value UUC (µl)	Observed Value (µl)	Error (µl)	Related Uncertainty on C.F.(k=2) at approx 95% Confidence Level ±
10	10.20	-0.20	0.13 µl
50	50.38	-0.38	0.13 µl
100	100.62	-0.62	0.13 µl

\*\*\*\*\* END OF REPORT \*\*\*\*\*

- NOTE:
- The result are valid only for this Calibrated item.
  - The calibration report shall not be reproduced without the written approval of calibration laboratory.
  - Result reported are valid at the time of & under the stated conditions of measurement
  - Laboratory standards are traceable to National/International Standards.
  - Calibration Certificates issued for Weight & Measure parameters i.e. Mass, Balance, Volumetric equipment, Measuring Scales / Tapes etc. are for scientific purpose only and should not be used for Trade / Commercial use.

CALIBRATED BY  
  
(CALIBRATION ENGINEER)

APPROVED BY  
  
ANWAR A. SIDDIQUI