

Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092
 Ph.: 011-42657237 Mob.: 9810804449, 8368671530
 E- mail: primac21@gmail.com



mat No.
(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC233921000000665F	Calibration Done On : 23.01.2021	S.R.F.No. & Date : 0001/22 & 22.01.2021
Issue Date : 23.01.2021	Calibration Valid Upto : 22.01.2022	Page : 1 of 1
UUC Receipt Date : --		

Calibrated for M/S.
Dr. Das Pathlab
 A-28, Lower Ground Floor, Ghanshyam Dass
 Jain Marg, South Extension Part-2,
 New Delhi-49

Identification on UUC (Unit Under Calibration)			
Instrument Name	Dig. Timer	Size/Range	0 to 60 Min
Least Count	1 Sec	Condition of Item	Good
Manufactured By	--	Calibration Performed at	Site
Manufacturing M/Sr No.	--	Location	Lab
Party ID Mark	DT-01	Ref. Standard	IS:1248

Calibration Procedure / Method : As per Calibration Procedure Document No. PCS/CPET/03, Section No.CPET-S-10

Environmental Conditions : Temperature : 25 ± 10°C. Relative Humidity : 50 ± 20 % RH

Standard & Major Equipment(s) Used :

Standard Equipments	Range /Serial No.	Calibrating Agency	Traceability Cert. No.	Certificate Date	Valid Up to
Digital Timer Calibrator	1 sec to 99999.9 Sec & 34452101	Transcal	CC2231200000 22263F	20.03.2020	19.03.2021

Calibration Results :

UUC Value (in Sec.)	Measured Value on Std. (in Sec.)	Deviation (in Sec.)	Related Uncertainty on C.F.(k=2) at approx 95% Confidence Level ±
300	300.07	-0.07	0.7 Sec.
600	600.13	-0.13	0.7 Sec.
1800	1800.2	-0.2	2.1 Sec.
3600	3600.3	-0.3	2.1 Sec.

***** END OF REPORT *****

NOTE:

- The result are valid only for this Calibrated item.
- The calibration report shall not be reproduced without the written approval of calibration laboratory.
- Result reported are valid at the time of & under the stated conditions of measurement
- Laboratory standards are traceable to National/International Standards.
- Calibration Certificates issued for Weight & Measure parameters i.e. Mass, Balance, Volumetric equipment, Measuring Scales / Tapes etc. are for scientific purpose only and should not be used for Trade / Commercial use.

CALIBRATED BY

(CALIBRATION ENGINEER)

APPROVED BY

ANWAR A. SIDDIQUI
(Q.M.T.M.)