

LAB MONTHLY SUMMARY



Lab Name **K M LAB COMPUTERIZED
MICRO LAB** Lab No **11217**
Month **October** Year **2021**
Constituent Group **Chemistry II**



Date of Result Entered : 20/10/2021

Date of Report Published : 09/11/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	910	168.35	10.59	17.82	211.5 mg/dl	2.42	1.18
2	UREA I	Urease Berthelot	Any Analyser	284	104.04	17.45	18.16	111.4 mg/dl	0.41	2.15
3	CREATININE I	Jaffes End point	Any Analyser	358	4.52	13.11	0.59	4.2 mg/dl	-0.54	0.06
4	URIC ACID I	Enzymatic / Uricase Colorimetric	Any Analyser	517	4.49	18.56	0.83	3.7 mg/dl	-0.95	0.07
5	CHOLESTEROL I	CHOD-PAP	Any Analyser	771	101.13	13.52	13.68	111 mg/dl	0.72	0.98
6	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	735	180.95	13.56	24.53	180.3 mg/dl	-0.03	1.81
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****