


12/2/21, 8:01 PM

LAB MONTHLY SUMMARY



Lab Name: DR DHAM S DIAGNOSTIC CENTRE  
 Lab No: 11089  
 Month: October  
 Year: 2021  
 Constituent Group: Chemistry II



Date of Result Entered : 21/10/2021

Date of Report Published : 01/11/2021

Sl. No.	Analyte	Method / Principle Name	Analysed Name	No of Participants	Participants			Mean Value	SD	SI
					CV	CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	154	100.15	10.14	11.07	116.6 mg/dl	0.67	1.78
2	UREA II	Urease UV / GLDH II	Mindray	124	26.65	11.74	3.13	13.3 mg/dl	-4.27	0.98
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	95	1.33	11.04	0.15	1.33 mg/dl	0.00	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	131	7.14	11.17	0.80	6 mg/dl	-1.43	0.14
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	138	5.16	8.55	0.44	4.7 g/dl	-1.04	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	137	3.16	8.45	0.27	2.2 g/dl	-3.60	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	130	5.84	13.64	0.81	6.3 mg/dl	0.57	0.14
8	CHOLESTROL II	CHOD-PAP II	Mindray	143	104.86	9.87	10.35	132 mg/dl	2.62	1.73
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	116	99.58	12.21	12.15	84.6 mg/dl	-1.23	2.26

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*

# Dr. Dham's Diagnostic Centre


A Specialist Centre

3856/23 Reghar Pura, Arya Samaj Road, Karol Bagh,  
New Delhi-110005 Ph.: 011-25852890, 25852891

The cuvettes of Mindray - 240 Biochemistry fully  
automated analyser has been replaced on  
2/12/2021.

Senior Lab Technician

Dr. GURMIT KAUR  
MBBS, D. PATH (DMC No. 3825)  
Consultant Biochemist

  
Dr. JASPREET KAUR  
MBBS, MD (DMC No. 15396)  
Consultant Biochemist

Ph. : 011-25852890, 25852891, E-mail : [dr.kawaldeepdham123@gmail.com](mailto:dr.kawaldeepdham123@gmail.com)

(REPORTS ARE NOT FOR MEDICO LEGAL PURPOSE, IT IS ONLY A PROFESSIONAL OPINION)

SERVICE REPORT

CUSTOMER DETAILS:



Name:	Dr. Dham's Diagnostic Centre		
Address:	3856/23 Reghari Puria Karol Bagh New Delhi - 110005		
Department:	Laboratory		
Contact Person:	Dr. Jaspreet Kaur		
Ph No:	Mobile No: 9810532989	E-Mail: j.kaur17@yahoo.com	


INSTRUMENT DETAILS:

Model Name:	BS-240	SLNo:	YX-64000027
Installation Date:	07/06/2018	Warranty Period:	
Status (RRC / PRC / Sold):		Offers:	
Software Version:			

Engineer's Remarks: Cuvettes Replaced on 2/12/21

Customer's Remarks: Satisfactory Results.

 SHIVAM GUPTA Service Engineer	 Customer's Signature / Name / Designation :
---	--



**Dr. Dham's Diagnostic Centre**  
A Specialist Centre

3856/23 Reghar Pura, Arya Samaj Road, Karol Bagh,  
New Delhi-110005 Ph.: 011-25852890, 25852891

**CAPA for the EQAS send in OCTOBER-2021**

The parameters Urea , Albumin, Cholesterol were recalibrated.

QC were within normal range.

Raw data is attached.

ILC was done from NABL accredited lab GAGAN.


The reports are attached along.

- The lamp was replaced on 25/10/21.  
The service report is attached along.

DATED - 06/11/2021

Senior Lab Technician

Dr. GURMIT KAUR  
MBBS, D. PATH (DMC No. 3825)  
Consultant Biochemist

  
Dr. JASPREET KAUR  
MBBS, MD (DMC No. 15396)  
Consultant Biochemist

Ph. : 011-25852890, 25852891, E-mail : dr.kawaldeepdham123@gmail.com

(REPORTS ARE NOT FOR MEDICO LEGAL PURPOSE IT IS ONLY A PROFESSIONAL OPINION)

SERVICE REPORT

CUSTOMER DETAILS:


Name:	Dr. Dham's Diagnostic Centre		
Address:	3856/23 Reghar Pura Koral Bagh New Delhi-110005		
Department:	LABORATORY		
Contact Person:	DR. JASPREET KAUR		
Ph No:	Mobile No: 9810532989	E-Mail: j.kaur70@yahoo.com	

INSTRUMENT DETAILS:

Model Name: BS-240	Sl.No: YX-64000027
Installation Date: 07/06/2018	Warranty Period: One Year upto 24/7/22
Status (RRC / PRC / Sold):	Offers:
Software Version:	

Engineer's Remarks: LAMP replaced on 25/10/21

Customer's Remarks: Satisfactory results

SHRUTI Gupta Mob-8808995541 Service Engineer	
Engineer's Signature / Name / Designation:	Customer's Signature / Name / Designation:

11/20/21, 1:50 PM

HOST

1637395425067.jpg

Admin

11/19 03:34 PM

Level-Derivings

Cumulative Sum

Twin-Plot

Results

Summary

QC Setup

QC Date 2021/11/06

2021/11/06

2021/11/06

2021/11/06

2021/11/06

971

971

971

971

971

971

971

971

971

971

971

GLUC

URIA

SGPT

SGOT

CHOL

PEP

TG

Co F

TP

ALB

BILITEST

BILITEST

Sample Type Serum

Sample Type Serum

Sample Type Serum

Sample Type Serum

Sample Type Serum

Sample Type Serum

Sample Type Serum

Sample Type Serum

Lot No.

Lot No.

Lot No.

Lot No.

Lot No.

Lot No.

Lot No.

Lot No.

Lot No.

Lot No.

Result

Result

Result

Result

Result

Result

Result

Result

Result

Result

217.1

100.3

128.6

145.3

251.7

79.09

173.53

13.5

9.05

5.56

6.11

1.61

Unit

Unit

Unit

Unit

Unit

Unit

Unit

Unit

262.31

98.09

144.1

146.76

264.10

86.51

170.43

9.12

5.07

0.41

1.12

4.94

9.61

9.70

13.21

5.77

8.52

0.46

0.25

0.41

1.12

2.78

2.78

2.78

2.78

2.78

2.78

2.78

2.78

2.78

2021/11/06 12:15:22

2021/11/06 12:10:16

2021/11/06 12:11:45

2021/11/06 12:12:03

2021/11/06 12:13:27

2021/11/06 12:19:15

2021/11/06 12:15:03

2021/11/06 12:14:45

2021/11/06 12:19:51

2021/11/06 12:16:33

2021/11/06 12:21:21

2021/11/06 12:21:30

Print P7

Archive P6

Compare P5

Print P4

Sort P3

Chems P2

Search P1

Exit

REDMI NOTE 10

LS communication error

**Dr. Dham's Diagnostic Centre**  
A Specialist Centre

3856/23 Reghar Pura, Arya Samaj Road, Karol Bagh,  
New Delhi-110005 Ph.: 011-25852890, 25852891

Date 06/11/2021  
Name MS. SHERRIN  
Ref. By SELF

Srl No. 29  
Age 19 Yrs.  
Address


Ref. No  
Sex F

Test Name	Value	Unit	Normal Value
<u>BIOCHEMISTRY</u>			
BLOOD UREA	21.6	mg/dl	10 - 50
ALBUMIN	4.6	gm/dl	3.4 - 4.8
TOTAL CHOLESTEROL	164.00	mg/dL	123.0 - 199.0

\*\*\*\* End Of Report \*\*\*\*

Senior Lab Technician

Dr. GURMIT KAUR  
MBBS, D. PATH (DMC No. 3625)  
Consultant Biochemist

  
Dr. JASPREET KAUR  
MBBS, MD (DMC No. 15396)  
Consultant Biochemist

Ph. : 011-25852890, 25852891, E-mail : dr.kawaldeepdham123@gmail.com

(REPORTS ARE NOT FOR MEDICO LEGAL PURPOSE, IT IS ONLY A PROFESSIONAL OPINION)



# GAGAN PATHOLOGY CENTRE

MRI, CT, ULTRASOUND, ECHO, TMT, EEG, EMG  
MAMMOGRAPHY, NCV, DEXA SCAN, OPG, ECG  
PFT, BERA, VEP, DIGITAL & PORTABLE X-RAYS

NABL Lab & NABH Accredited CT/MRI Imaging

Date 06/11/2021

Name MS. SHERRIN

Srl No. 30

Ref. No

Age 19 Yrs.

Sex F

Ref. By DR.DHAM

Address

Test Name	Value	Unit	Normal Value
<u>BIOCHEMISTRY</u>			
BLOOD UREA	22.0	mg /dl	10 - 50
ALBUMIN	4.8	gm/dl	3.4 - 4.8
TOTAL CHOLESTEROL	162.00	mg/dL	123.0 - 199.0

\*\*\*\* End Of Report \*\*\*\*

**Dr. R.K. Garg**  
MBBS, MD  
Senior Consultant Pathologist

Page 1 of 1

**Dr. Ankur Garg**  
MBBS, MD  
Consultant Pathologist

# NABL Accredited ; Tests without # are out of Scope.

Expert and Reliable Opinion for over 25 Years

H.O. : F-26/21-22, Sector-7, Rohini, Delhi-110085 Email : gaganlab@yahoo.com website : www.gaganlab.com

On Panel : CGHS, D.G.E.H.S., DJB, NDPL, ECHS, DDA, ITPO, NDMC, MCD, LIC (TPA), ESI, DELHI UNIVERSITY



LAB MONTHLY SUMMARY

Lab Name DR DHAM S DIAGNOSTIC CENTRE  
 Lab No 11089  
 Month September  
 Year 2021  
 Constituent Group Chemistry II

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/09/2021

Date of Report Published : 05/10/2021

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	SD	Your Value	gDI	U
1	GLUCOSE II	GOD-POD II	Mindray	143	403.99	11.17	45.12	490 mg/dl	1.91	7.55
2	UREA II	Urease UV / GLDH II	Mindray	112	91.16	8.16	7.44	94 mg/dl	0.38	1.41
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	87	3.35	10.93	0.37	2.4 mg/dl	-2.60	0.08
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	125	2.12	13.90	0.30	2.3 mg/dl	0.61	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	121	5.13	7.50	0.38	5.5 g/dl	0.96	0.07
6	ALBUMIN II	BCG - Colorimetric II	Mindray	124	3.21	8.83	0.28	2.2 g/dl	-3.56	0.05
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	122	8.39	12.38	1.04	9.9 mg/dl	1.45	0.19
8	CHOLESTROL II	CHOD-PAP II	Mindray	138	113.82	10.26	11.68	120 mg/dl	0.53	1.99
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	126	230.68	9.14	21.09	220 mg/dl	-0.51	3.76

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

*Dr. Dham's Diagnostic Centre*

A Specialist Centre

3856/23 Reghar Pura, Arya Samaj Road, Karol Bagh,  
New Delhi-110005 Ph.: 011-25852890, 25852891

**CAPA for the EQAS Send in SEPTEMBER-2021**

The parameters Creatinine, Albumin, were recalibrated.

QC were within normal range.


Raw data is attached.

ILC was done from NABL accredited lab GAGAN.

The reports are attached along.

Senior Lab Technician

Dr. GURMIT KAUR  
MBBS, D. PATH (DMC No. 3825)  
Consultant Biochemist

  
Dr. JASPREET KAUR  
MBBS, MD (DMC No. 15396)  
Consultant Biochemist

Ph. : 011-25852890, 25852891, E-mail : [dr.kawaldeepdham123@gmail.com](mailto:dr.kawaldeepdham123@gmail.com)

(REPORTS ARE NOT FOR MEDICO LEGAL PURPOSE, IT IS ONLY A PROFESSIONAL OPINION)

11/23/21, 4:55 PM

163766456719.jpg

Program Result QC Utility Alarm Exit

11/23/21 4:55 PM

Control: REDMI P 971    QC Date: 2021/10/05    Summary: 2021/10/05

Control	Lot No.	Chem	Sample Type	Lot No.	Result	Unit	Set Mean	Std SD	Flag	Run Date
NEED1 P	971	CHD	Serum	3025.6		mg/dL	264.18	13.21	OK	2021/10/05
NEED1 P	971	CHD	Serum	2560.8		mg/dL	170.43	6.57	OK	2021/10/05
NEED1 P	971	TC	Serum	163.66		g/dL	9.17	0.46	OK	2021/10/05
NEED1 P	971	T.P	Serum	11.08		g/dL	9.17	0.46	OK	2021/10/05
NEED1 P	971	T.P	Serum	10.87		g/dL	9.17	0.46	OK	2021/10/05
NEED1 P	971	ALB	Serum	5.37		g/dL	5.02	0.25	OK	2021/10/05
NEED1 P	971	BILI T	Serum	4.41		mg/dL	4.97	0.41	OK	2021/10/05
NEED1 P	971	BILI D	Serum	1.29		mg/dL	1.59	0.13	OK	2021/10/05
NEED1 P	971	ALP	Serum	523.9		U/L	425.27	30.02	OK	2021/10/05
NEED1 P	971	URE AVT	Serum	12.0		mg/dL	12.17	0.61	OK	2021/10/05
NEED1 P	971	CRET AVT	Serum	2.69		mg/dL	3.48	0.244	OK	2021/10/05
NEED1 P	971	CRET AVT	Serum	3.20		mg/dL	3.48	0.244	OK	2021/10/05

Search F1    Chems F2    Stat F3    Peak Curve F4    Summary F5    Archive F6    Print F7

REDMI NOTE 10

https://mail.google.com/mail/u/0/?tab=rm&ogblinbox?projector=1

**Dr. Dham's Diagnostic Centre**  
A Specialist Centre

3856/23 Reghar Pura, Arya Samaj Road, Karol Bagh,  
New Delhi-110005 Ph.: 011-25852890, 25852891


Date	05/10/2021	Srl No.	69	Ref. No	
Name	MR. OM PRAKASH	Age	60 Yrs.	Sex	M
Ref. By	SELF	Address			
Test Name		Value	Unit	Normal Value	

Test Name	Value	Unit	Normal Value
<u>BIOCHEMISTRY</u>			
SERUM CREATININE	2.10	mg/dl	0.6 - 1.4
ALBUMIN	3.6	gm/dl	3.4 - 4.8

\*\*\*\* End Of Report \*\*\*\*

Senior Lab Technician

Dr. GURMIT KAUR  
MBBS, D. PATH (DMC No. 3825)  
Consultant Biochemist

  
Dr. JASPREET KAUR  
MBBS, MD (DMC No. 15396)  
Consultant Biochemist



The Association of Practising Pathologists (Regd.)

# GAGAN PATHOLOGY CENTRE

MRI, CT, ULTRASOUND, ECHO, TMT, EEG, EMG  
MAMMOGRAPHY, NCV, DEXA SCAN, OPG, ECG  
PFT, BERA, VEP, DIGITAL & PORTABLE X-RAYS


**NABL Lab & NABL Accredited CT/MRI Imaging**      **Sri No. 70**      **Ref. No**  
**Name** MR. OM PRAKASH      **Age** 60 Yrs.      **Sex** M  
**Ref. By** DR.DHAM      **Address**

Test Name	Value	Unit	Normal Value
<u>BIOCHEMISTRY</u>			
SERUM CREATININE	2.30	mg/dl	0.6 - 1.4
ALBUMIN	3.7	gm/dl	3.4 - 4.8

\*\*\*\* End Of Report \*\*\*\*

**Dr. R.K. Garg**  
 MBBS, MD  
 Senior Consultant Pathologist

Page 1 of 1

  
**Dr. Ankur Garg**  
 MBBS, MD  
 Consultant Pathologist


Expert and Reliable Opinion for over 25 Years

# NABL Accredited ; Tests without # are out of Scope.  
 H.O. : F-26/21-22, Sector-7, Rohini, Delhi-110085 Email : gaganlab@yahoo.com website : www.gaganlab.com

On Panel : CGHS, D.G.E.H.S., DJB, NDPL, ECHS, DDA, ITPO, NDMC, MCD, LIC (TPA), ESI, DELHI UNIVERSITY

External Quality Assurance Scheme - Print Monthly Summary

Lab Name DR DHAM S DIAGNOSTIC CENTRE Lab No 11089  
 Month June Year 2021  
 Constituent Group Chemistry II



Date of Result Entered : 09/06/2021

Date of Report Published : 03/07/2021

Sl. No.	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Yours Value	SDI	SI
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	146	197.88	10.47	20.72	217 mg/dl	0.92	3.43
2	UREA II	Urease UV / GLDH II	Mindray	114	77.16	9.25	7.14	114.5 mg/dl	5.23	1.34
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	83	4.35	10.51	0.46	5.42 mg/dl	2.34	0.10
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Mindray	129	3.86	15.41	0.59	4.31 mg/dl	0.76	0.10
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	125	5.31	8.52	0.45	6.44 g/dl	2.50	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	130	3.29	10.15	0.33	3.37 g/dl	0.24	0.06
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	115	4.59	22.58	1.04	4.6 mg/dl	0.01	0.19
8	CHOLESTROL II	CHOD-PAP II	Mindray	134	107.35	10.66	11.44	127 mg/dl	1.72	1.98
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	128	139.58	10.12	14.12	150 mg/dl	0.76	2.50

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*

## *Dr. Dham's Diagnostic Centre*

A Specialist Centre

3856/23 Reghar Pura, Arya Samaj Road, Karol Bagh,  
New Delhi-110005 Ph.: 011-25852890, 25852891

### **CAPA for the EQAS send in JULY-2021**

The parameters Urea , Creatinine, protein  
were recalibrated.

QC were within normal range.

Raw data is attached.

ILC was done from NABL accredited lab  
GAGAN.

The reports are attached along.

Senior Lab Technician

Dr. GURMIT KAUR  
MBBS, D. PATH (DMC No. 3825)  
Consultant Biochemist

Dr. JASPREET KAUR  
MBBS, MD (DMC No. 15396)  
Consultant Biochemist

Ph. : 011-25852890, 25852891, E-mail : [dr.kawaldeepdham123@gmail.com](mailto:dr.kawaldeepdham123@gmail.com)

(REPORTS ARE NOT FOR MEDICO LEGAL PURPOSE, IT IS ONLY A PROFESSIONAL OPINION)

LAB P. 971      QC Date: 2021/07/05      2021/07/05

Lot No.	Chem	Sample Type	Lot No.	Result	Unit	Set Mean	Set SD	Flag	Plan Date/Time
971	GLUC	Serum		271.2	mg/dL	262.31	13.12		2021/07/05 01:3
971	GLUC	Serum		293.6	mg/dL	262.31	13.12	DEL	2021/07/05 01:3
971	UREA	Serum		107.1	mg/dL	98.89	4.94		2021/07/05 12:5
971	SCPT	Serum		150.0	U/L	144.1	9.61		2021/07/05 01:3
971	SGOT	Serum		133.3	U/L	146.76	9.78		2021/07/05 12:5
971	CHO	Serum		295.9	mg/dL	264.18	13.21		2021/07/05 12:5
971	HDL	Serum		67.86	mg/dL	66.51	5.77		2021/07/05 01:0
971	TG	Serum		182.07	mg/dL	170.43	8.52		2021/07/05 12:5
971	Ca +	Serum		11.0	mg/dL	10.61	0.53		2021/07/05 01:0
971	T.P	Serum		9.68	g/dL	9.12	0.46		2021/07/05 12:5
971	ALB	Serum		4.71	g/dL	5.02	0.25		2021/07/05 12:5
971	CREA-J	Serum		3.51	mg/dL	3.48	0.244		2021/07/05 01:0

Search F1      Chemis F2      Sort F3      Peak Curve F4      Comment F5      Archive F6      Print F7



**Dr. Dham's Diagnostic Centre**  
A Specialist Centre

3856/23 Reghar Pura, Arya Samaj Road, Karol Bagh,  
New Delhi-110005 Ph.: 011-25852890, 25852891


Date 05/07/2021 Sri No. 10 Ref. No  
Name MR. SHER SINGH Age 57 Yrs. Sex M  
Ref. By DR. ARVIND BANSAL Address

Test Name	Value	Unit	Normal Value
<b>KIDNEY FUNCTION TEST (RFT)</b>			
BLOOD UREA	44.0	mg /dl	10 - 50
SERUM CREATININE	0.90	mg\dl	0.6 - 1.4
TOTAL PROTEIN	7.4	gm/dl	6.6 - 8.3

\*\*\*\* End Of Report \*\*\*\*

Senior Lab Technician

Dr. GURMIT KAUR  
MBBS, D. PATH (DMC No. 3825)  
Consultant Biochemist

  
Dr. JASPREET KAUR  
MBBS, MD (DMC No. 15396)  
Consultant Biochemist

Ph. : 011-25852890, 25852891. E-mail : dr.kawaldeepdham123@gmail.com

**GAN PATHOLOGY & IMAGING PVT. LTD.**  
(Previously Gagan Pathology Centre)

NABH (MIS) Accredited CT / MRI Imaging

MRI, CT, ULTRASOUND, ECHO, TMT, EEG, EMG  
MAMMOGRAPHY, NCV, DEXA SCAN, OPG, ECG  
PFT, BERA, VEP, DIGITAL & PORTABLE X-RAYS

Date 05/07/2021  
Name MR. SHER SINCH  
Ref. By DR.DHAM

**Lab Report**  
Sl No. 49  
Age 57 Yrs.  
Address


Ref. No  
Sex M

Test Name	Value	Unit	Normal Value
<b>BIOCHEMISTRY</b>			
BLOOD UREA	44.9	mg /dl	10 - 50
SERUM CREATININE	0.98	mg/dl	0.6 - 1.4
TOTAL PROTEIN	7.8	gm/dl	6.6 - 8.3

\*\*\*\* End Of Report \*\*\*\*

**Dr. R.K. Garg**  
MBBS, MD  
Senior Consultant Pathologist

Page 1 of 1

  
**Dr. Ankur Garg**  
MBBS, MD  
Consultant Pathologist

Expert and Reliable Opinion for over 30 Years

H.O. : F-26/21-22, Sector-7, Rohini, Delhi-110085 Email : gaganlab@yahoo.com website : www.gaganlab.com

On Panel : CGHS, D.G.E.H.S., DJB, NDPL, ECHS, DDA, ITPO, NDMC, MCD, LIC (TPA), ESI, DELHI UNIVERSITY

VIEW LAB MONTHLY SUMMARY

Lab Name DR DHAM S DIAGNOSTIC CENTRE  
 Lab No 11089  
 Month May  
 Year 2021  
 Constituent Group Chemistry II

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 18/05/2021

Date of Report Published : 05/06/2021


Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Mindray	129	133.59	8.67	11.58	130 mg/dl	-0.31	2.04
2	UREA II	Urease UV / GLDH II	Mindray	107	64.54	11.57	7.47	73 mg/dl	1.13	1.44
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	82	1.81	9.34	0.17	2 mg/dl	1.12	0.04
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Mindray	110	2.52	17.62	0.44	2.34 mg/dl	-0.40	0.08
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	120	5.42	11.27	0.61	5.84 g/dl	0.69	0.11
6	ALBUMIN II	BCG - Colorimetric II	Mindray	119	3.36	11.48	0.39	2.87 g/dl	-1.27	0.07
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	116	6.89	14.60	1.01	6.8 mg/dl	-0.09	0.19
8	CHOLESTROL II	CHOD-PAP II	Mindray	136	112.55	11.60	13.06	137 mg/dl	1.87	2.24
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	123	223.20	9.92	22.13	257 mg/dl	1.53	3.99

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

Lab Name DR DHAM S DIAGNOSTIC CENTRE Lab No 11089  
 Month July Year 2021  
 Constituent Group Chemistry II



Date of Result Entered : 16/07/2021

Date of Report Published : 02/08/2021

S. No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Year Value	SDI	U
						CV	SD	SDI			
1	GLUCOSE II	GOD-POD II	Mindray	142	151.78	7.48	11.36	163 mg/dl	0.99	1.91	
2	UREA II	Urease UV / GLDH II	Mindray	119	31.60	13.28	4.20	38 mg/dl	1.53	0.77	
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	94	2.80	10.45	0.29	3.25 mg/dl	1.54	0.06	
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	125	2.92	12.35	0.36	2.5 mg/dl	-1.16	0.06	
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	125	5.39	8.15	0.44	5.9 g/dl	1.16	0.08	
6	ALBUMIN II	BCG - Colorimetric II	Mindray	135	3.35	9.01	0.30	3.25 g/dl	-0.33	0.05	
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	130	4.85	19.12	0.93	5.3 mg/dl	0.48	0.16	
8	CHOLESTROL II	CHOD-PAP II	Mindray	144	113.36	10.13	11.48	129.5 mg/dl	1.41	1.91	
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	136	175.27	10.64	18.64	188.7 mg/dl	0.72	3.20	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*

Lab Name **DR DHAM S DIAGNOSTIC CENTRE** Lab No **11089**  
 Month **August** Year **2021**  
 Constituent Group **Chemistry II**



Date of Result Entered : 18/08/2021

Date of Report Published : 01/09/2021

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SDI	U
						CV	SD				
1	GLUCOSE II	GOD-POD II	Mindray	131	257.89	9.56	24.66	294 mg/dl	1.46	4.31	
2	UREA II	Urease UV / GLDH II	Mindray	112	44.17	9.62	4.25	47 mg/dl	0.67	0.80	
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	88	1.87	14.43	0.27	2.1 mg/dl	0.85	0.06	
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	118	4.15	10.91	0.45	4.2 mg/dl	0.11	0.08	
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	129	5.35	8.73	0.47	5.4 g/dl	0.11	0.08	
6	ALBUMIN II	BCG - Colorimetric II	Mindray	127	3.33	9.26	0.31	3.3 g/dl	-0.10	0.05	
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	121	3.42	24.47	0.84	3.7 mg/dl	0.33	0.15	
8	CHOLESTROL II	CHOD-PAP II	Mindray	133	110.62	9.25	10.24	128 mg/dl	1.70	1.78	
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	135	133.53	15.27	20.39	140 mg/dl	0.32	3.51	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential  
 Contact details:  
 Email: clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*

PDF Created Using



# Camera Scanner

Easily Scan documents & Generate PDF



<https://play.google.com/store/apps/details?id=photo.pdf maker>