



## CALIBRATION CERTIFICATE

ULR NO- CC325721000003027F

M/S:- Dr.Reetu's Path Lab Bachpan Clinical, Mohna Road, Ballabgarh	Service request No & date	KL/214-06/09/2021
	Date of receiving	06/12/2021
	Date of calibration	06/12/2021
	Suggested Due Date	05/12/2022
	Date of issue	06/12/2021

Instrument Details			
Instrument Name	Digital Temp. Indicator Fitted With Freezer	S. No.	-----
Make	-----	Location	Lab
Range / Size	-8°C to -2°C	I.D. No.	DRPL/DTI-02
Least Count	0.1 °C	Visual Inspection	OK

Detail of reference standards & Major equipments used					
Standards	Make	ID No.	Certificate No	Next Due Date	Traceability
RTD with Indicator	SBI/YCT	RTD-01	CC205321000000086F	31/01/2022	KRC

Environmental Condition	Temperature	(25±4) °C	Calibration Reference	ITS:1990
	Relative Humidity	(50±20 )%	Calibration Procedure	WI(T):05

### Calibration Results

Serial No.	Nominal Reading (DUC) in (°C)	Observed Reading in (°C)	Uncertainty At 95% C.L. (coverage factor k=2)
01.	-8.2	-8.12	±0.35 °C
02.	-6.4	-6.25	
03.	-4.2	-4.63	
04.	-2.3	-2.51	

### Remarks:

- This result of calibration refers to only to the particular item submitted for calibration.
- The above results are valid at the time of and under the stated conditions measurement.
- This report should not be reproduced except in full without the written permission of CEO.
- The traceability used in calibration is traceable to national / international standards.

Calibrated By  
(Calibration Engg./TM)  
Page 1 of 1



Approved By  
(QM/TM)  
(C.S.MAHAPATRA)

---End of Report---

Format No. QF:47, issue 01, revision 01