

LAB MONTHLY SUMMARY



Lab Name **CPL DIAGNOSTICS PVT LTD** Lab No **14825**
 Month **July** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 16/07/2021

Date of Report Published : 02/08/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	986	147.82	9.98	14.75	183.33 mg/dl	2.41	0.94
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	567	32.30	14.53	4.69	29.3 mg/dl	-0.64	0.39
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	720	2.75	11.18	0.31	3.24 mg/dl	1.59	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	775	2.96	15.26	0.45	2.91 mg/dl	-0.11	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	742	5.53	9.90	0.55	5.73 g/dl	0.36	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	711	3.31	9.70	0.32	3.62 g/dl	0.97	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	834	5.17	19.21	0.99	5.66 mg/dl	0.49	0.07
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	984	115.76	13.00	15.04	110.5 mg/dl	-0.35	0.96
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	818	177.61	12.70	22.56	192.5 mg/dl	0.66	1.58

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****