

LAB MONTHLY SUMMARY



Lab Name **CPL DIAGNOSTICS PVT LTD** Lab No **14825**
 Month **September** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 20/09/2021

Date of Report Published : 05/10/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1006	386.50	11.89	45.96	480 mg/dl	2.03	2.90
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	576	89.70	12.21	10.96	85.7 mg/dl	-0.37	0.91
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	718	3.30	11.92	0.39	4.02 mg/dl	1.83	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	830	2.22	17.99	0.40	2.24 mg/dl	0.05	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	751	5.37	9.93	0.53	4.18 g/dl	-2.23	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	733	3.19	9.97	0.32	3.28 g/dl	0.28	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	799	8.43	13.92	1.17	10.5 mg/dl	1.76	0.08
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	973	115.65	11.78	13.62	135.5 mg/dl	1.46	0.87
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	806	223.24	11.93	26.62	256.2 mg/dl	1.24	1.88

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****