

LAB MONTHLY SUMMARY



Lab Name **CPL DIAGNOSTICS PVT LTD** Lab No **14825**
 Month **November** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 22/11/2021

Date of Report Published : 03/12/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1004	144.70	10.23	14.80	164 mg/dl	1.30	0.93
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	577	42.79	12.32	5.27	29.38 mg/dl	-2.54	0.44
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	709	9.02	12.78	1.15	11.05 mg/dl	1.76	0.09
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	940	2.43	23.94	0.58	2.46 mg/dl	0.05	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	768	5.44	10.95	0.60	5.64 g/dl	0.34	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	760	3.22	9.92	0.32	3.28 g/dl	0.19	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	801	8.10	14.01	1.14	10.12 mg/dl	1.78	0.08
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	993	112.61	12.13	13.66	126.25 mg/dl	1.00	0.87
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	843	156.72	12.90	20.21	187.5 mg/dl	1.52	1.39

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****