

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE



Customer Name & Address:	M/S, Indian Medical Diagnostic Centre, 593/A1, KJS Complex, Near Canara Bank Main Road, New Bus Stand, Bhavani, Erode -638301	Certificate No:	AMS/LAB/289/0112
		Issuing Date:	19/11/2021
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Digital Thermometer	Id. No:	IMDC/LAB/DT/001
Make:	R tek	Department:	Laboratory
Model:	RT - 116	Calibrated On:	18/11/2021
Sl. No:	---	Calibration Due Suggested:	17/11/2022
Type:	PT 100	DUC Condition on Receipt:	OK
Range & Least Count:	0 - 6°C / LC 1°C		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	PC/01/21/ATL/0107-01	22/01/2022
2	Thermo Hygrometer	Beetech/088	PC/01/21/ATL/0194-01	22/01/2022

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

### Test Results:

SL.No.	Refrigerator Place	DUC Reading (°C)	Standard Reading (°C)	Correction (°C)
1	Freezer	-1.56	-1.49	+0.07
2		-1.42	-1.33	+0.09
3	Chiller	1.39	1.46	+0.07
4		1.51	1.58	+0.07
5	Chamber I	2.55	2.61	+0.06
6		2.38	2.48	+0.10
7	Chamber II	3.18	3.26	+0.08
8		3.32	3.38	+0.06
9	Chamber III	4.27	4.34	+0.07
10		4.39	4.49	+0.10

Allowable Tolerance: User Defined.

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading is taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:  
(Calibration Engineer)

*B. S. S.*



Approved by:  
(Technical Manager)

*B. S. S.*