

## VIEW LAB MONTHLY SUMMARY

| Lab Name  | ANNAI DIAGNOSTIC LAB            |   | <a href="#">Details About Robust Analysis</a> |                    |                       |                 |       |              |       |      |
|---|---------------------------------|---|---|--------------------|-----------------------|-----------------|-------|--------------|-------|------|
| Lab No  | 13696                           |   | <a href="#">Detail About Monthly Summary</a>  |                    |                       |                 |       |              |       |      |
| Month   | August                          |   | <a href="#">Detail about SDI</a>              |                    |                       |                 |       |              |       |      |
| Year  | 2021                            |   |   |                    |                       |                 |       |              |       |      |
| Constituent Group   | Chemistry II                    |   |   |                    |                       |                 |       |              |       |      |
| <a href="#">Click on the analyte to view Graphical Data</a> |                                 |   | <a href="#">All Analyser Result</a>           |                    | <a href="#">Print</a> |                 |       |              |       |      |
|   |                                 |   | Date of Result Entered : 20/08/2021           |                    |                       |                 |       |              |       |      |
|   |                                 |   | Date of Report Published : 01/09/2021         |                    |                       |                 |       |              |       |      |
| Sl.No   | Analyte                         | Method / Principle Name                         | Analyser Name                                 | No of Participants | DV                    | Participants CV | SD    | Your Value   | SDI   | U    |
| 1   | <a href="#">GLUCOSE II</a>      | GOD-POD II                                      | Any Analyser (Automation / Semi Automation)   | 918                | 247.66                | 10.88           | 26.94 | 228.61 mg/dl | -0.71 | 1.78 |
| 2   | <a href="#">UREA II</a>         | Others ( Urease Berthelot / Ned Dye ) II        | Any Analyser (Automation / Semi Automation)   | 873                | 44.63                 | 13.73           | 6.13  | 50.78 mg/dl  | 1.00  | 0.41 |
| 3   | <a href="#">CREATININE II</a>   | Jaffes Kinetic-Alkaline Picrate II              | Any Analyser (Automation / Semi Automation)   | 702                | 1.90                  | 13.28           | 0.25  | 1.86 mg/dl   | -0.16 | 0.02 |
| 4   | <a href="#">T.BILIRUBIN II</a>  | Diazonium Salt ( Colorimetric ) / Jendrassik II | Any Analyser (Automation / Semi Automation)   | 741                | 4.12                  | 14.43           | 0.60  | 4.15 mg/dl   | 0.05  | 0.04 |
| 5   | <a href="#">T-PROTEIN II</a>    | Biuret - Colorimetric II                        | Any Analyser (Automation / Semi Automation)   | 727                | 5.45                  | 10.82           | 0.59  | 5.85 g/dl    | 0.68  | 0.04 |
| 6   | <a href="#">ALBUMIN II</a>      | BCG - Colorimetric II                           | Any Analyser (Automation / Semi Automation)   | 672                | 3.29                  | 10.23           | 0.34  | 2.86 g/dl    | -1.28 | 0.03 |
| 7   | <a href="#">URIC ACID II</a>    | Enzymatic / Uricase Colorimetric II             | Any Analyser (Automation / Semi Automation)   | 764                | 3.93                  | 24.71           | 0.97  | 2.97 mg/dl   | -0.99 | 0.07 |
| 8   | <a href="#">CHOLESTROL II</a>   | CHOD-PAP II                                     | Any Analyser (Automation / Semi Automation)   | 895                | 112.13                | 11.42           | 12.81 | 106.09 mg/dl | -0.47 | 0.86 |
| 9   | <a href="#">TRIGLYCERIDE II</a> | GPO-PAP / Enzymatic Colorimetric / End Point II | Any Analyser (Automation / Semi Automation)   | 784                | 130.89                | 14.50           | 18.98 | 145.38 mg/dl | 0.76  | 1.36 |
| <b>SDI Range</b>  |                                 | <b>Interpretation</b>                           |   |                    |                       |                 |       |              |       |      |
| Within -1.00 to +1.00                                       |                                 | Excellent.                                      |   |                    |                       |                 |       |              |       |      |
| Within ±1.01 to ±2.00                                       |                                 | Good.   |   |                    |                       |                 |       |              |       |      |
| Within ±2.01 to ±2.99                                       |                                 | Accept with caution. Warning Signal.            |   |                    |                       |                 |       |              |       |      |
| Beyond ±3.0   |                                 | Unacceptable performance. Action Signal.        |   |                    |                       |                 |       |              |       |      |

## Links