

VIEW LAB MONTHLY SUMMARY

Lab Name	ANNAI DIAGNOSTIC LAB		Details About Robust Analysis							
Lab No	13696		Detail About Monthly Summary							
Month	October		Detail about SDI							
Year	2021									
Constituent Group	Chemistry II									
Click on the analyte to view Graphical Data			All Analyser Result		Print					
			Date of Result Entered : 20/10/2021							
			Date of Report Published : 01/11/2021							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1038	107.73	10.13	10.92	101 mg/dl	-0.62	0.68
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	887	29.20	15.70	4.58	32.3 mg/dl	0.68	0.31
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	744	1.32	12.58	0.17	1.47 mg/dl	0.90	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	801	7.05	13.48	0.95	7.28 mg/dl	0.24	0.07
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	793	5.35	10.66	0.57	6.19 g/dl	1.47	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	769	3.19	10.00	0.32	2.95 g/dl	-0.75	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	833	5.90	15.69	0.93	4.7 mg/dl	-1.30	0.06
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	908	106.08	11.36	12.05	103 mg/dl	-0.26	0.80
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	754	101.65	15.55	15.81	120 mg/dl	1.16	1.15
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links