

LAB MONTHLY SUMMARY



Lab Name **ANNAI DIAGNOSTIC LAB** Lab No **13696**
 Month **November** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 11/11/2021

Date of Report Published : 03/12/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1004	144.70	10.23	14.80	141.44 mg/dl	-0.22	0.93
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	844	43.73	12.85	5.62	48.01 mg/dl	0.76	0.39
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	709	9.02	12.78	1.15	7.48 mg/dl	-1.34	0.09
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	940	2.43	23.94	0.58	2.66 mg/dl	0.40	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	768	5.44	10.95	0.60	5.61 g/dl	0.29	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	760	3.22	9.92	0.32	2.91 g/dl	-0.97	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	801	8.10	14.01	1.14	6.86 mg/dl	-1.09	0.08
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	993	112.61	12.13	13.66	119.73 mg/dl	0.52	0.87
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	843	156.72	12.90	20.21	126.89 mg/dl	-1.48	1.39

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****