

VIEW LAB MONTHLY SUMMARY

Lab Name	NARAYAN PATHOLOGY AND BIOPSY CENTRE				Details About Robust Analysis					
Lab No	11024				Detail About Monthly Summary					
Month	November				Detail about SDI					
Year	2021									
Constituent Group	Chemistry II									
Click on the analyte to view Graphical Data				All Analyser Result		Print				
				Date of Result Entered : 13/11/2021						
				Date of Report Published : 03/12/2021						
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Biosystems	275	145.00	8.96	12.99	138 mg/dl	-0.54	1.57
2	UREA II	Urease UV / GLDH II	Biosystems	209	41.88	11.84	4.96	30 mg/dl	-2.40	0.69
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Biosystems	215	9.38	9.85	0.92	8.47 mg/dl	-0.98	0.13
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Biosystems	234	2.58	18.61	0.48	2.67 mg/dl	0.19	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	237	5.33	8.82	0.47	6.02 g/dl	1.47	0.06
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	241	3.21	9.83	0.32	3.4 g/dl	0.60	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Biosystems	225	8.31	9.48	0.79	6.41 mg/dl	-2.41	0.11
8	CHOLESTROL II	CHOD-PAP II	Biosystems	259	112.50	9.93	11.17	101 mg/dl	-1.03	1.39
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Biosystems	229	156.47	10.68	16.72	158 mg/dl	0.09	2.21
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

