



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF CENTRIFUGE

Calibration Certificate No.	21900002063	Issued On:	25/06/2021
ULR No.	CC214621900002063F	Calibration Date:	22/06/2021
Job Identification No.	ICS/C/AS/06/2063	Next Calibration Date:	21/12/2021
Ref. No.	SRF, Dated - 18/06/2021		

CALIBRATED FOR:
M/s Polo Labs Pvt Ltd.,
C/o IVY Hospital,
Khanna, Opp. Khanna City Centre.

EQUIPMENT DESCRIPTION

Name	Centrifuge, Fitted with Digital RPM Indicator		
Make/Model	REMI / R-8C Plus	Visual Inspection	Ok
Serial No.	ZELN-41549	I.D. No.	IVY/KH/CF/01
Range	Up to 6000 RPM	Least Count	100 RPM
Location	---	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

STANDARD REFERENCE DETAILS Traceable to National Standard (NPL, New Delhi)

Name of the Standard	Make	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
Digital Tachometer	Lutron/ ICS/SR/INS/02	21900000002	Indian Calibration Services, Delhi	01-01-2021	01-01-2022

USED EQUIPMENT DETAILS	Not Applicable
PRINCIPLE/METHODOLOGY OF CALIBRATION:	As per Calibration Procedure No.: ICS/CAL/SOP-RPM01 (Comparison Method)

RESULTS: Mechanical Calibration
MEASUREMENT: (in RPM)

CENTRIFUGE' READING (AVERAGE)	CORRECTED VALUE OF TACHOMETER READING (AVG.)	UNCERTAINTY AT APPROX 95% C.L AND COVERAGE FACTOR K=2
1800	1803	± 63
2900	2906	± 63
3600	3610	± 64
4900	4915	± 64
5800	5919	± 65

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

***END OF REPORT ***

Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.

2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.

3. Total liability of our Organisation is limited to the invoiced amount.

4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.

5. In case any reconfirmation of contents of this Calibration Certificate is required, Please contact our office.

6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

Authorised Signatory