



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF MICROPIPETTE

Calibration Certificate No.	21700006924	Issued On:	25/06/2021
ULR No.	CC214621700006924F	Calibration Date:	22/06/2021
Job Identification No.	ICS/C/MVD/06/6924	Next Calibration Date:	21/12/2021
Ref. No.	SRF, Dated - 18/06/2021		

CALIBRATED FOR:

M/s Polo Labs Pvt Ltd.,
C/o IVY Hospital,
Khanna, Opp. Khanna City Centre.

EQUIPMENT DESCRIPTION

Name	Variable Micropipette		
Make/Model	Erba	Visual Inspection	Ok
Serial No.	YE4A237193	I.D. No.	MP-02
Range	100 µl - 1000 µl	Least Count	5 µl
Location	Lab	Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0 °C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD EQUIPMENT DETAILS Traceable to National Standards

Name	Make	Certificate No.	Calibration Agency	Calibration Date	Valid Upto
Stainless Steel Wire Weights	Weightronics	WMCL/E/2020-05/3364	WMCL	31/05/2020	30/05/2023
RTD Sensor with Digital Indicator	Yudian/ICS/STD/01	ETL/2070/21/C/F/060	FCRI, Palakkad	24-25/02/2021	25/02/2022

USED EQUIPMENT DETAILS: Calibrated Balance, Distilled Water & Glass wares

PRINCIPLE/METHODOLOGY OF CALIBRATION: As per Calibration Procedure No.: ICS/CAL/SOP-M03 (Gravimetric Method), ISO 8655-6

RESULTS: Mechanical Calibration
Below Volume is determined at 27°C

U.U.C. Reading (in ml)	Standard Measured Volume (in ml) (Average)	Uncertainty at approx 95% C.L. and coverage factor k=2
0.1000 (0100 µl)	0.10049 (0100.49 µl)	± 3 µl
0.5000 (0500 µl)	0.50074 (0500.74 µl)	± 3 µl
1.0000 (1000 µl)	1.00093 (1000.93 µl)	± 3 µl

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

END OF REPORT

Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.

2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.

3. Total liability of our Organisation is limited to the invoiced amount.

4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.

5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.

6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

Authorised Signatory



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
Phone / Fax : (011) 22453259, 22434362, 9599482307
E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF MICROPIPETTE

Calibration Certificate No.	21700006925	Issued On:	25/06/2021
ULR No.	CC214621700006925F	Calibration Date:	22/06/2021
Job Identification No.	ICS/C/MVD/06/6925	Next Calibration Date:	21/12/2021
Ref. No.	SRF, Dated - 18/06/2021		

CALIBRATED FOR:

M/s Polo Labs Pvt Ltd.,
C/o IVY Hospital,
Khanna, Opp. Khanna City Centre.

EQUIPMENT DESCRIPTION

Name	Variable Micropipette		
Make/Model	Erba		
Serial No.	YE182AG0203261	Visual Inspection	Ok
Range	10 µl - 100 µl	I.D. No.	MP-01
Location	Lab	Least Count	1 µl
		Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS

Temperature

25.0 °C ± 3.0° C

Humidity

50 ± 10 % RH

STANDARD EQUIPMENT DETAILS Traceable to National Standards

Name	Make	Certificate No.	Calibration Agency	Calibration Date	Valid Upto
Stainless Steel Wire Weights	Weightronics	WMCL/E/2020-05/3364	WMCL	31/05/2020	30/05/2023
RTD Sensor with Digital Indicator	Yudian/ICS/STD/01	ETL/2070/21/C/F/060	FCRI, Palakkad	24-25/02/2021	25/02/2022

USED EQUIPMENT DETAILS

Calibrated Balance, Distilled Water & Glass wares

PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS/CAL/SOP-M03 (Gravimetric Method), ISO 8655-6

RESULTS: Mechanical Calibration Below Volume is determined at 27°C

U.U.C. Reading (in ml)	Standard Measured Volume (in ml) (Average)	Uncertainty at approx 95% C.L and coverage factor k=2
0.0100 (010 µl)	0.009541 (009.541 µl)	± 0.6 µl
0.0500 (050 µl)	0.048468 (048.468 µl)	± 0.6 µl
0.1000 (100 µl)	0.097984 (097.984 µl)	± 0.6 µl

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.
***END OF REPORT ***

- Note: 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

Authorised Signatory



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
 Phone / Fax : (011) 22453259, 22434362, 9599482307
 E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF MICROPIPETTE

Calibration Certificate No.	21700006967	Issued On:	25/06/2021
ULR No.	CC214621700006967F	Calibration Date:	23/06/2021
Job Identification No.	ICS/C/MVD/06/6967	Next Calibration Date:	22/12/2021
Ref. No.	SRF, Dated - 18/06/2021		

CALIBRATED FOR: M/s Polo Labs Pvt Ltd.,
 C/o IVY Hospital,
 Khanna, Opp. Khanna City Centre.

EQUIPMENT DESCRIPTION

Name	Variable Micropipette		
Make/Model	Precica		
Serial No.	---	Visual Inspection	Ok
Range	10 µl - 100 µl	I.D. No.	MP-05
Location	Lab	Least Count	1 µl
		Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD EQUIPMENT DETAILS Traceable to National Standards

Name	Make	Certificate No.	Calibration Agency	Calibration Date	Valid Upto
Stainless Steel Wire Weights	Weightronics	WMCL/E/2020-05/3364	WMCL	31/05/2020	30/05/2023
RTD Sensor with Digital Indicator	Yudian/ICS/STD/01	ETL/2070/21/C/F/060	FCRI, Palakkad	24-25/02/2021	25/02/2022

USED EQUIPMENT DETAILS

Calibrated Balance, Distilled Water & Glass wares

PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS/CAL/SOP-M03 (Gravimetric Method), ISO 8655-6

RESULTS: Mechanical Calibration
 Below Volume is determined at 27°C

U.U.C. Reading (in ml)	Standard Measured Volume (in ml) (Average)	Uncertainty at approx 95% C.L and coverage factor k=2
0.0100 (010 µl)	0.010086 (010.086 µl)	± 0.6 µl
0.0500 (050 µl)	0.050173 (050.173 µl)	± 0.6 µl
0.1000 (100 µl)	0.100315 (100.315 µl)	± 0.6 µl

REMARKS:The recommended date for next calibration is mentioned, as asked by the customer.

***END OF REPORT ***

- Note :**
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
 - Total liability of our Organisation is limited to the invoiced amount.
 - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
 - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
 - The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
 Indian Calibration Services

Checked by

Authorized Signatory



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF MICROPIPETTE

Calibration Certificate No.	21700006966	Issued On:	25/06/2021
ULR No.	CC214621700006966F	Calibration Date:	23/06/2021
Job Identification No.	ICS/C/MVD/06/6966	Next Calibration Date:	22/12/2021
Ref. No.	SRF, Dated - 18/06/2021		

CALIBRATED FOR:

M/s Polo Labs Pvt Ltd.,
C/o IVY Hospital,
Khanna, Opp. Khanna City Centre.

EQUIPMENT DESCRIPTION

Name	Variable Micropipette		
Make/Model	---	Visual Inspection	Ok
Serial No.	---	I.D. No.	MP-04
Range	10 µl - 100 µl	Least Count	1 µl
Location	Lab	Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD EQUIPMENT DETAILS

Traceable to National Standards

Name	Make	Certificate No.	Calibration Agency	Calibration Date	Valid Upto
Stainless Steel Wire Weights	Weightronics	WMCL/E/2020-05/3364	WMCL	31/05/2020	30/05/2023
RTD Sensor with Digital Indicator	Yudian/ICS/STD/01	ETL/2070/21/C/F/060	FCRI, Palakkad	24-25/02/2021	25/02/2022

USED EQUIPMENT DETAILS

Calibrated Balance, Distilled Water & Glass wares

PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS/CAL/SOP-M03
(Gravimetric Method), ISO 8655-6

RESULTS: Mechanical Calibration

Below Volume is determined at 27°C

U.U.C. Reading (in ml)	Standard Measured Volume (in ml) (Average)	Uncertainty at approx 95% C.L and coverage factor k=2
0.0100 (010 µl)	0.009936 (009.936 µl)	± 0.6 µl
0.0500 (050 µl)	0.049873 (049.873 µl)	± 0.6 µl
0.1000 (100 µl)	0.099822 (099.822 µl)	± 0.6 µl

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

***END OF REPORT ***

- Note: 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

Authorised Signatory