

CALIBRATION CERTIFICATE

Name of the customer	Polo lab Pvt. ltd khanna
Location	Khanna
Make	Beckman Coulter
Product Name	DXH500
Serial No	AZ080390

This is to certify that the cell counter installed at the above mentioned customer site has been calibrated with (Lot no-492113600 dt- 5-March-2021)

S.No	Procedure	Parameter	Remarks
1	Startup	WBC/RBC/HGB/MCV/RDW/PLT/MPV	PASS
2	Reproducibility	WBC/RBC/HGg/MCV/RDW/PLT/MPV	PASS
3	QCTri Level	WBC/RBC/HGB/MCV/RDW/PLT/MPV	PASS
4	Calibration	WBC/RBC/HGB/MCV/ PLT/MPV	PASS

Date of Calibration: 01-MARCH-2021	Due Date of Calibration: 01-MARCH-2022
------------------------------------	--

For Able Diagnostic Systems Pvt. Ltd




Ajay Mahajan
General – Manager operations