

Certificate No: CC2471



S.M. ENTERPRISES


Lab / Office : Plot No. 12, Sanjeevaiah Nagar, Co-operative Society,
Besides Tadband Hanuman Temple, Sikh Village, Secunderabad - 500 009, T. S. INDIA.
Phone : 040-27840697 | 48541008, Telefax : 040-27840798,
Cell : 99482 93008 | 99482 91008 | 98480 20798 | 99482 92008,
E-mail : msudheer_98@smentp.org | msudheer_98@yahoo.com
Website : www.smentp.org

CALIBRATION CERTIFICATE

SME/QM/FF/43	ULR NO : CC24712100000811F		Page : 1 of 2
Certificate No.:	SME/PH/TH/697	Date of Issue :	20-Nov-2021
Date of Calibration :	20-Nov-2021	Recom. Due Dt :	19-Nov-2022
Name and Address of the Customer :		Customer Reference :	
M/s. PACE HOSPITALS Hitech City , T.S.		SRF No :	091
		Calibrated at :	On-Site
		Date of Receipt :	20-Nov-2021
		Condition of item :	GOOD
		Field of Calibration :	Thermal
Details of Instrument: (UUC / DUC)			
Description :	Incubator	Make :	--
Range :	37 °C	Model :	--
Accuracy :	± 1 °C	ID No :	PHM/BME/LAB/INCUB/01
Resolution :	0.1 °C	Location :	4TH FLOOR
Master Insurment Details :			
Name	ID No.	Valid Upto:	Traceability
Digital Temperature indicator with Sensor	SME/DTI/010	Dt: 12/12/2020 DueDt:12/12/2021	C-201205-17-2
Work Instruction No :		SME/QM/WI/02	
Calibration Environment :			
Temperature :	24.3 °C	Relative Humidity :	40.3 % Rh

CALIBRATION CERTIFICATE

Checked By:  M. SUJEET KUMAR Technical-Manager	Authorised Signatory  SUDHEER M. KUMAR / M. SUJEET KUMAR QM & MD / D.QM & TM
--	---





Certificate No: CC2471

S.M. ENTERPRISES

Lab / Office : Plot No. 12, Sanjeevaiah Nagar, Co-operative Society,
Besides Tadbund Hanuman Temple, Sikh Village, Secunderabad - 500 009, T. S. INDIA.
Phone : 040- 27840697 | 48541008, Telefax : 040-27840798,
Cell : 99482 93008 | 99482 91008 | 98480 20798 | 99482 92008,
E-mail : msudheer_98@smentp.org | msudheer_98@yahoo.com
Website : www.smentp.org

Certificate No.: SME/PH/TH/697 Page : 2 of 2



CALIBRATION RESULT


Temperature				
Sr. No	Standard Reading °C	Test Reading °C	Error (±) °C	Uncertainty (±) °C
1	37.41	37.0	0.41	2.80

Note:

- The above UUC/DUC was Calibrated at Lab.
- Definition of UUC / DUC : Unit under Calibration / Device under Calibration.
- The reported Expanded Uncertainty is calculated at 95.45 % with covarger factor k = 2.
- Test reading given was the average of 5 Measurements.
- The report shall not be reproduced in full/part without prior permission of S.M Enterprises.
- The report refers only to particular item(s) submitted for calibration.
- The calibration results reported are valid at the time of and under the stated conditions of the measurements.
- The Equipment was Calibrated without any adjustment.
- The above Expanded Uncertainty includes Repeatability, Stability, Resolution & Uncertainty of Measurements.
- S.M Enterprises is not liable for any change in calibration data & performance specifications on account of malfunctioning of Standard / instrument / Equipments covered by this report due to damage caused to it after issuance of this report.
- All Precautions have been taken for any error or omission while calibrating the instruments and issuing this certificate. However S.M Enterprises shall not be liable for any loss or liability that may be arised to any party in this regards.
- The Equipments user for Calibration were Calibrated and traceable to National / International Standards.

END OF CERTIFICATE

<p>Checked By:</p>  <p>M. SUJEET KUMAR Technical Manager</p>	<p>Authorised Signatory</p>  <p>SUDHEER M. KUMAR / M. SUJEET KUMAR QM & MD / D.QM & TM</p>
---	--



CALIBRATION CERTIFICATE