

Issued from : Choksi Laboratories Limited  
 Survey No. 9/1, Balaji Tulsiyana Industrial Estate,  
 Kumed, Off. M.R. 10 Toll Naka, Indore -452010 (M.P.), INDIA

## CERTIFICATE OF CALIBRATION



**CHOKSI**  
 LABORATORIES LIMITED

CALIBRATION CERTIFICATE NUMBER :03900003/9/21-22

EQUIPMENT NAME :-

MICROPIPETTE

RECEIVED ON	CALIBRATED ON	CALIBRATION DUE DATE RECOMMENDED	PAGE NO.
28.05.2021	01.06.2021	01.06.2022	1 of 1

I. Calibrated for:

PRESTIGE MEDICAL INVESTIGATION, INDORE  
 12/2, NEW AGRAWAL NAGAR, AGRASEN SQUARE  
 Indore, Madhya Pradesh  
 India - 452001

II. Description and Identification of Instrument :

Sample Id :	03900003/9
Make :	Labsystems
Model :	Finnpipette
Type :	Not Mentioned
Serial No. :	G 38497

Capacity :	5 µl
Calibrated Range :	5 µl
Least Count :	Not applicable
Condition On Receipt :	Satisfactory
Identification No.:	Not Mentioned

III. Environment Condition :-

Temperature :	20 ± 2 °C
Relative Humidity :	50 ± 10 % RH

IV. Standard/Calibrator and Measuring Equipment Used :

Nomenclature	Range	Uncertainty	Certificate No.	Traceability	Validity
Weight Box(E1)	1 mg to 200 g	± 0.001 to 0.03 mg	SCW-1120-268	CC-2813	05.11.2023
With Precession Balance, Distilled Water & Digital Thermometer					

V. Results: (a) Volume

Serial No.	Scale	UUC <sup>#</sup>	Standard Average	Correction	Uncertainty (For k=2, 95.45 % C.L.) : ±
1	µl	5.0	5.0239275	0.023927	0.054

Note : 01. UUC<sup>#</sup> : Unit Under Calibration

: 02. Certificate shall not be reproduced in full without the written approval of the laboratory

Other Details

01. Certificate issue date : 02/06/2021
02. Location of Calibration : Lab
03. Temperature of Distilled water used for calibration is 20.4 °C.
04. The results relate only to the item calibrated

VI. Inference:

01. The correction is positive except when indicated other wise It has to be algebraically added to the UUC<sup>#</sup> reading in order to get the actual reading .

VII. Reference:

01. The Calibration is carried out as per the CLL Calibration Procedure  
 No. CLL-IDR-SWP-CAL-000045.
02. All the values have been rounded off as per IS:2-1960.
03. The laboratory is NABL Accredited in the field of Mechanical as per Certificate No. , CC-2089, ULR NO.:CC208921000001239F

VIII. Details of Adjustment made ( If any)


Not Done

CALIBRATED BY IDR01183

AUTHORISED SIGNATORY

Signature Not Verified  
 Mesh J. Dure  
 02.06.2021 15:06

\*\*END OF CALIBRATION CERTIFICATE\*\*



CI-I-MSP-201-Annex.-02  
Calibration Tag

Ver 5.0 Effective Date: 01-04-2015



Last Ver 4.0 Dated 01-04-2017

Next Review Due On 01-04-2023

Equipment ID:                     

COC ID:                     

Equipment Name:                     

Calibrated On:                       
(Date of Initiation)

Calibration Completed on:                       
(Date of completion)

Recommended Due Date:                     

Valid Calibration Range:                     

Equipment Serial No.:                     

Signature:                     

NOTE: Calibration Tag (to be displayed on the instrument after calibration, along with Equipment Tag, CI-I-MSP-5.5-01-05).  
Replace after every calibration:- Attach the old calibration tags with its corresponding COC (Certificate of Calibration).  
Give COC as follows:

CI - 3 Character Branch Code - 'Coc' - 3 Character Equipment ID / ddmmyyy  
(For E.g. IDR-Coc-001/010415 represents calibration tag or Equipment ID '001' at CI's Indore Branch, on 01 APRIL 2015).