

Issued from : Choksi Laboratories Limited
 Survey No. 9/1, Balaji Tulsiyana Industrial Estate,
 Kumedl, Off. M.R. 10 Toll Naka, Indore -452010 (M.P.), INDIA

CERTIFICATE OF CALIBRATION



CALIBRATION CERTIFICATE NUMBER :03900003/10/21-22

EQUIPMENT NAME :- **MICROPIPETTE**

RECEIVED ON	CALIBRATED ON	CALIBRATION DUE DATE RECOMMENDED	PAGE NO.
28.05.2021	01.06.2021	01.06.2022	1 of 1

I. Calibrated for :
 PRESTIGE MEDICAL INVESTIGATION, INDORE
 12/2, NEW AGRAWAL NAGAR, AGRASEN SQUARE
 Indore, Madhya Pradesh
 India - 452001

II. Description and Identification of Instrument :

Sample Id :	03900003/10
Make :	Thermo SCIENTIFIC
Model :	Finnpipette
Type :	FINNPIPETTE F3
Serial No. :	GJ30634

Capacity :	10 µl
Calibrated Range :	10 µl
Least Count :	Not applicable
Condition On Receipt :	Satisfactory
Identification No.:	Not Mentioned

III. Environment Condition :-

Temperature :	20 ± 2 °C
Relative Humidity :	50 ± 10 % RH

IV. Standard/Calibrator and Measuring Equipment Used :

Nomenclature	Range	Uncertainty	Certificate No.	Traceability	Validity
Weight Box(E1)	1 mg to 200 g	± 0.001 to 0.03 mg	SCW-1120-268	CC-2813	05.11.2023
With Precession Balance, Distilled Water & Digital Thermometer					

V. Results: (a) Volume

Serial No.	Scale	UUC #	Standard Average	Correction	Uncertainty (For k=2, 95.45 % C.L.) : ±
1	µl	10.0	10.0661066	0.066107	0.054

- Note :** 01. UUC # : Unit Under Calibration
 : 02. Certificate shall not be reproduced in full without the written approval of the laboratory
- Other Details** 01. Certificate issue date : 02/06/2021
 02. Location of Calibration : Lab
 03. Temperature of Distilled water used for calibration is 20.3 °C.
 04. The results relate only to the item calibrated

VI. Inference:

01. The correction is positive except when indicated other wise It has to be algebraically added to the UUC # reading in order to get the actual reading .

VII. Reference:

01. The Calibration is carried out as per the CLL Calibration Procedure
 No. CLL-IDR-SWP-CAL-000045.
 02. All the values have been rounded off as per IS:2-1960.
 03. The laboratory is NABL Accredited in the field of Mechanical as per Certificate No. , CC-2089. ULR NO.:CC208921000001240F

VIII. Details of Adjustment made (If any)

Not Done

CALIBRATED BY IDR01183

AUTHORISED SIGNATORY

Signature Not Verified
 Mitesh J. Doshi IDR01183
 02.06.2021 15:46

END OF CALIBRATION CERTIFICATE

5. (1) - Annex.-02

Calibration Tag

Ver 5.0 Effective Date: 2019

LAB Ver 4.0 Dated 01-04-2017

Next Review Due On: 01-04-2022



Equipment ID: Not mentioned

COC ID: 03900003/19/21-22

Equipment Name: micropipette

Calibrated On: 01/06/2021
(Date of Initiation)

Calibration Completed on: 01/06/2021
(Date of completion)

Recommended Due Date: 01/06/2022

Valid Calibration Range: 10 µg

Equipment Serial No.: 9530634

Signature: [Signature]

NOTE: Calibration Tag (to be displayed on the instrument after calibration, along with Equipment Tag, CLL-MSP-5-5-01-05).
Replace after every calibration. Attach the old calibration tags with its corresponding CoC (Certificate of Calibration).
Give CoC as follows:

CLL - 3 Character Branch Code - 'Coc' - 3 Character Equipment ID / ddmmyyy
(For E.g. IDR-Coc-001/010415 represents calibration tag of Equipment ID '001' at CLL's Indore Branch, on 01 APRIL 2015).

