

CERTIFICATE OF CALIBRATION

Issued from : Choksi Laboratories Limited
 Survey No. 87, Balaji Tulyehana Industrial Estate,
 Kumedhi, Ch. N.R. 10 Toll Naka,
 Indore - 452018 (M.P.), INDIA



CALIBRATION CERTIFICATE NUMBER: 03900003-07/21-23

EQUIPMENT NAME : Fully Automatic Biochemistry Analyzer(Temperature)		CALIBRATION DUE DATE RECOMMENDED	PAGE NO.
RECEIVED ON	CALIBRATED ON		
28.05.2021	31.05.2021	31.05.2022	1 of 1

I. Calibrated for :
 PRESTIGE MEDICAL INVESTIGATION, ROOM#
 122, NEW AGRAWAL NAGAR, AGRASEN SQUARE
 Indore, Madhya Pradesh
 India - 452001

II. Description and Identification of Instrument :

Sample Id:	03900003-07
Make:	ABM
Model:	ADM-200
Type:	SUN6020 1655T
Serial No.	Not mentioned

Range:	30.0 to 45.0 °C
Calibrated Range:	30.1 to 45.1 °C
Least Count:	0.1 °C
Condition on receipt:	Satisfactory
Identification No. :	Not mentioned

III. Environment Condition-

Temperature :	25 ± 5 °C
Relative Humidity :	45 to 75 % RH

IV. Standard/Calibrator and Measuring Equipment Used :

Nomenclature	Range	Uncertainty	Certificate No.	Traceability	Validity
Digital thermometer with sensor	-79.8 to 299.8 °C	± 0.142 to 0.350 °C	M-200629-21-1	CC-2634	16.07.2021

V. Results: (Temperature)

Thermal Calibration					
Serial No.	Unit	UUC* Average	Standard Average	Correction	Uncertainty (For k=2, 95.45 % C.L.) : ±
1	°C	30.06	30.44	0.38	0.82
2	°C	35.08	35.02	-0.06	0.82
3	°C	37.04	37.06	0.02	0.82
4	°C	40.06	40.70	0.64	0.82
5	°C	45.08	45.10	0.02	0.82

Note : 01.UUC* : Unit Under Calibration

- : 02. Certificate shall not be reproduced in full without the written approval of the laboratory
- : 03. The results relate only to the item calibrated.

Other Details

- 01. Certificate issue date : 02/06/2021
- 02. Location of calibration: Site
- 03. The results relate only to the item calibrated

VI. Inference:

01. The correction is positive except when indicated other wise it has to be algebraically added to the UUC* reading in order to get the actual reading .

VII. Reference:

- 01. The Calibration is carried out as per the CLL Calibration Procedure
 No. CLL-IDR-SWP-CAL-00041
- 02. All the values have been rounded off as per IS-2-1960.
- 03. The laboratory is NABL Accredited having certificate No. CC-2089. ULR NO. CC208921000001237F

VIII. Details of Adjustment Made (if any)

Not Done

CALIBRATED BY IDR00910

AUTHORISED SIGNATORY

Signature Not Verified
 Nilesh J. Doshi 18101151
 02.06.2021 12:44

END OF CALIBRATION CERTIFICATE

CLL-MSP-5.5-01-Annex.-02
Calibration Tag



Ver 5.0 Effective Date: 01-04-2019

Last Ver 4.0 Dated 01-04-2017

Next Review Due On: 01-04-2022

Equipment ID: Not Mentioned
CoC ID: 03900003-02/21-22
Equipment Name: Fully Automatic Bioche
Calibrated On: (Date of Initiation) 31/05/2021
Calibration Completed on: (Date of completion) 31/05/2021
Recommended Due Date : 31/05/2022
Valid Calibration Range : Refer CoC
Equipment Serial No.: SUN602016SST
Signature: *[Signature]*

NOTE: Calibration Tag (to be displayed on the instrument after calibration, along-with Equipment Tag, CLL-MSP-5.5-01-05).
Replace after every calibration. Attach the old calibration tags with its corresponding CoC (Certificate of Calibration).
Give CoC as follows:
CLL - 3 Character Branch Code - 'CoC' - 3 Character Equipment ID / ddmmyyyy
(For E.g. IDR-CoC-001/010415 represents calibration tag or Equipment ID '001' at CLL's Indore Branch, on 01 APRIL 2015).

