

CERTIFICATE OF CALIBRATION

Issued from : Chokai Laboratories Limited
 Survey No. 9/1, Balaji Tulsiyana Industrial Estate,
 Kumedl, Off. M.R. 10 Toll Naka,
 Indore -452010 (M.P.), INDIA



CHOKAI
 LABORATORIES LIMITED

CALIBRATION CERTIFICATE NUMBER :03900003-05/21-22

EQUIPMENT NAME : DIGITAL TIMER			
RECEIVED ON	CALIBRATED ON	CALIBRATION DUE DATE REQUESTED BY CUSTOMER	PAGE NO.
28.05.2021	31.05.2021	31.05.2022	1 of 1

I. Calibrated for : PRESTIGE MEDICAL INVESTIGATION, INDORE
 12/2, NEW AGRAWAL NAGAR, AGRASEN SQUARE
 Indore, Madhya Pradesh
 India - 452001

II. Description and Identification of Instrument :

Sample Id:	03900003-05
Make:	Not mentioned
Model:	Not mentioned
Type:	Digital
Serial No.:	Not mentioned

Range:	Not Mentioned
Calibrated Range:	1740 sec
Least Count:	0.1 sec
Condition On Receipt :	Satisfactory
Identification No. :	Not Mentioned

III. Calibration Environment :

Temperature:	25 ± 2 °C
Relative Humidity:	50 ± 10% RH

IV. Standard/Calibrator and Measuring Equipment Used :

Nomenclature	Make	Uncertainty	Certificate No.	Traceability	Validity
Digital Timer Interval Meter	Boltronics	±0.001100.577 Sec	C&U/CAL/20-07/051	CC-2216	16.07.2021

V. Results :

Serial No.	UUC Average Value (M)			Standard Average Value (S)	Correction (S-M)	Uncertainty (For ka2, 95.45 % C.L.) : ±
	Min.	Sec.	1/100 Sec.	Sec.	Sec.	Sec.
1	0.00	10.00	0.14	10.063	0.062	0.663
2	1.00	0.00	0.15	63.080	0.078	0.663
3	10.00	0.00	0.17	600.158	0.156	0.663
4	20.00	0.00	0.15	1200.212	0.210	0.664
5	30.00	0.00	0.15	1800.230	0.228	0.663

Note : 01. UUC^o = Unit Under Calibration

02. Calibration Certificate shall not be reproduced in full without the written approval of the laboratory

Other Details :

- 01. Certificate Issue Date : 02/06/2021
- 02. Location of Calibration Lab:
- 03. The results relate only to the item calibrated

VI. Inference:

01. The correction is positive except when indicated, other wise it has to be algebraically added to the UUC^o reading in order to get the actual reading.

VII. Reference

- 01. The Calibration is carried out as per the CLL Calibration Procedure No. CLL-IDR-SWP-CAL-00030
- 02. All the values have been rounded off as per IS 2-1960.
- 03. The laboratory is NABL Accredited having Certificate Number CC-2089, ULR NO. CC208921000001235F

VIII. Details of Adjustment made (if any)

Not Done

CALIBRATED BY IDR00910

AUTHORISED BY

Signature Not Verified
 Manish J. Goyal IDR01151
 02.06.2021 14:44

END OF CALIBRATION CERTIFICATE

CLL-MSP-5.5-01-Annex.-02
Calibration Tag



Ver 5.0 Effective From: 01-04-2019

Last Ver 4.0 Dated 01-04-2017

Next Review Due On: 01-04-2022

Equipment ID:	Not Mentioned
CoC ID:	03900003-04/21-22
Equipment Name:	Digital Thermo Hygrometer
Calibrated On: (Date of Initiation)	01/06/2021
Calibration Completed on: (Date of completion)	01/06/2021
Recommended Due Date :	01/06/2022
Valid Calibration Range :	Refer CoC
Equipment Serial No.:	Not mentioned
Signature:	[Signature]

NOTE: Calibration Tag (to be displayed on the instrument after calibration, along-with Equipment Tag, CLL-MSP-5.5-01-05).
Replace after every calibration. Attach the old calibration tags with its corresponding CoC (Certificate of Calibration).
Give CoC as follows:
CLL - 3 Character Branch Code - 'CoC' - 3 Character Equipment ID / ddmmyyy
(For E.g. IDR-CoC-001/010415 represents calibration tag or Equipment ID '001' at CLL's Indore Branch, on 01 APRIL 2015).



देव सर्वकार्येषु सा ॥



TEMPERATURE

27.9°C
AM 12:32 55%

CLOCK / HUMIDITY

HTC-1