

VIEW LAB MONTHLY SUMMARY

Lab Name	SHREE PATH LAB	Details About Robust Analysis
Lab No	14718	Detail About Monthly Summary
Month	October	Detail about SDI
Year	2021	
Constituent Group	Chemistry II	

[Click on the analyte to view Graphical Data](#)
[All Analyser Result](#)
[Print](#)

Date of Result Entered : 18/10/2021

Date of Report Published : 01/11/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1038	107.73	10.13	10.92	159.1 mg/dl	4.7	0.68
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	612	28.11	15.98	4.49	29.1 mg/dl	0.22	0.36
3	CREATININE II	Enzymatic Colorimetric II	Any Analyser (Automation / Semi Automation)	91	1.34	11.76	0.16	1.35 mg/dl	0.06	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	801	7.05	13.48	0.95	6.74 mg/dl	-0.33	0.07
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	793	5.35	10.66	0.57	5.74 g/dl	0.68	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	636	3.21	9.77	0.31	3.9 g/dl	2.20	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	833	5.90	15.69	0.93	7.28 mg/dl	1.49	0.06
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	908	106.08	11.36	12.05	143 mg/dl	3.06	0.80
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	754	101.65	15.55	15.81	72.9 mg/dl	-1.82	1.15

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.

Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out