

APEX QUALITY CERTIFICATION SERVICES PVT. LTD.







CALIBRATION CERTIFICATE

CERTIFICATION	Page 01 of 01	
CERTIFICATE NO.:AQCS/21041302.10004	Date of Calibration : 13/04/2021	
	Recommended due date : 12/04/2022	
CUSTOMER'S DETAILS	INSTRUMENT DESCRIPTION	
M/S RELIABLE DIAGNOSTIC CENTER (Kota) C-14/2, Talwandi, Nursery Yojana, Kota, Rajasthan	Nomenclature : Centrifuge Machine Make/Model : Remi /R-303 Sr. No./Ident. No. : NA / RDCK/LAB/CFM/01 Range : 0-3500 Least Count : As per results	
Service Request No. : 210304/01 Dated: 04/03/2021	Location : Lab Cal. Performed At : Site	

STANDARD EQUIPMENT USED:

Nomenclature	Identification No.	Traceability Certificate No.	Valid up to
Digital Tachometer	S999205	NCL/B/2021/34/01	25/02/2022
5411 #B 6414 4515 4 6614 5 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1			

ENVIRONMENTAL CONDITIONS:

PROCEDURE FOLLOWED:

Temperature: 23 ± 5°C Relative Humidity: 50± 10% CP/RM-01

CALIBRATION RESULT

S. No.	Set position at DUC	Measured Value on Standard in RPM
1	1	591
2	2	1139
3	3	2352
4	4	2914
5	5	3479

REMARKS:

- 1. Expanded uncertainty of Measurement at approximately 95 % confidence level for coverage factor k = 2 is ± 2.37% of reading.
- The result reported in this certificate is valid at the time of and under the started conditions of Measurement for the particular item and instrument identified above.
- 3. This certificate shall not be reproduced, except in full, unless a written permission for the publication of an approved abstract has been obtained from the Director, Apex Quality Certification Services Pvt. Ltd. Jaipur

4. Format Number.: F-06/00

Validated by

(Sachin Yadav)

Approved by

(R. R. Yadav) Quality Manager