

## APEX QUALITY CERTIFICATION SERVICES PVT. LTD.

Add.: 161, Kasturba Nagar, Nirman Nagar, Jaipur-302 019 Ph.: 0141-4036776, 9414079483, 9636293333

Email: apex.qcs@gmail.com



NABL ACCREDITED LAB Cert. No.: CC-2298

CALIBRATION CERTIFICATE

ULR: CC229821000000941F	DISCIPLINE Thermal	Page 01 of 01
CUSTOMER'S DETAILS	CERTIFICATE NO.:	AQCS/21041302.001
M/S RELIABLE DIAGNOSTIC CENTER (Kota)	Date of Issue of Certificate	15/04/2021
C-14/2, Talwandi, Nursery Yojana,	Date of Calibration	13/04/2021
Kota, Rajasthan	Recommended due date	12/04/2022

Service Request No	0.	210408/03 08/04/2021		INSTRUMENT DESCRIPTION		
Service Request Da	ate			Nomenclature	Temperature Indicator with sensor of Refrigerator	
DUC Receipt Date		13/04/2021		Make	Kelvinator	
DUC Condition		ok		Model	KFL 234	
Cal. Performed At		Site		Sr. No. / Ident. No.	3904102613 / RDCK/LAB/REF/TI/01	
Procedure Followe	ed	CP/TH-08		Range	2 to 8	°C
Environmental Condition		Least Count	0.1	°C		
Temperature	25±5°C	Rel. Hum.	50±10%	Location	Lab	

## STANDARD EQUIPMENT USED:

Nomenclature	Range	Uncertainty (±)	Identification No.	Traceability Certificate No.	Valid up to
PT100, 4 wire RTD with Indicator	-80 to 200 °C	0.1 to 0.15 °C	11339 /16A046131	TL/020/319.2.1	28-05-2021

## **CALIBRATION RESULT**

	0.12.2.1.1.1.1.1.1.1.1.1.1.1.1.1.1.1.1.1		
Value on standard (°C)	Measurement DUC (°C)	Error (°C)	Expanded Uncertainty ± (°C)
2.02	2.2	0.18	2.5
2.90	3.1	0.20	2.5
4.88	5.1	0.22	2.5
6.65	6.9	0.25	2.5
7.52	7.8	0.28	2.5
7.52	7.8	0.28	2.3

## **REMARKS:**

- 1. Expanded uncertainty of Measurement at approximately 95 % confidence level for coverage factor k = 2 is as given above.
- 2. The result reported in this certificate is valid at the time of and under the stated conditions of Measurement for the particular item and instrument identified above.
- 3. This certificate shall not be reproduced, except in full, unless a written permission for the publication of an approved abstract has been obtained from the Director, Apex Quality Certification Services Pvt. Ltd. Jaipur

4. Format Number.: F-06/00

Calibrated by

(Sachin Yadav)

JAIPUR JAIPUR

Approved by

(R. R. Yadav) Quality Manager