

MINU CLINICAL LABORATORY (558)  
 Outlier And Analyte Summary Report  
 Result Details For Cycle No 210410 and Sample No 04

Report Date : 02/11/2021

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Neutrophil.	-	40-70%	40-70%
@ Lymphocyte.	-	25-50%	25-50%
@ Eosinophil.	-	1-6%	1-6%

Legend @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⌚ : Delayed Result Entry

Total Parameters	3
Not Evaluated Parameters	0
Evaluated Parameters	3
Outlier Parameters (X)	0
EQAS Score Haematology-Peripheral Smear Count	100.00 %

Qualitative

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416,  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroli Road, Kurla (W),  
 Mumbai - 400070

End of Report

MINU CLINICAL LABORATORY (558)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 210410 and Sample No 04

Report Date : 02/11/2021

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated 🕒 : Delayed Result Entry			

Problem Classification: \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Corrective Action: \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Reviewed by: \_\_\_\_\_

Dated: \_\_\_\_\_

MINU CLINICAL LABORATORY (558)

Outlier And Analyte Summary Report

Result Details For Cycle No 210410 and Sample No 01

Report Date : 02/11/2021

Instrument : Nihon Kohden Celltac Alpha

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	5.80	5.77	-0.10	0.02
@ Total Leucocytes (WBC) count	/c.mm	6480.00	6,390.00	0.21	0.68
@ Haemoglobin	g/dL	14.70	14.65	-0.11	-0.10
@ MCV (Mean Corpuscular Volume)	fL	87.70	85.05	0.01	-0.01
@ MCH (Mean Corpuscular Hb)	pg	25.20	25.30	-0.07	-0.16
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	28.80	29.85	-0.28	-0.14
@ Platelet Count	1000/ $\mu$ L	197.00	196.00	-0.68	-0.21
@ Hematocrit	%	51.10	49.20	-0.10	-0.15
@ RDW CV%	%	19.90	19.80	-1.28	-0.79

Legend @ : z score  $\leq$  2.0 - Acceptable  
 ! : 2.0 < z score < 3.0 - Warning  
 X : z score  $\geq$  3 - Unacceptable  
 # : Not Evaluated  
 ⌚ : Delayed Result Entry  
 \* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416,  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroli Road, Kurla (W),  
 Mumbai - 400070

End of Report

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable I : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

MINU CLINICAL LABORATORY (558)  
Outlier And Analyte Summary Report  
Result Details For Cycle No 210410 and Sample No 02

Report Date : 02/11/2021

Instrument : Nihon Kohden Celltac Alpha

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	4.60	4.56	0.13	0.02
@ Total Leucocytes (WBC) count	/c.mm	5290.00	4,795.00	1.14	0.68
@ Haemoglobin	g/dL	11.50	11.55	-0.09	-0.10
@ MCV (Mean Corpuscular Volume)	fL	86.40	84.15	-0.02	-0.01
@ MCH (Mean Corpuscular Hb)	pg	24.80	25.25	-0.24	-0.16
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	28.70	30.05	0.00	-0.14
@ Platelet Count	1000/ $\mu$ L	238.00	242.00	0.27	-0.21
@ Hematocrit	%	39.90	38.45	-0.20	-0.15
@ RDW CV%	%	19.90	19.70	-0.29	-0.79

Legend @ : z score  $\leq$  2.0 - Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score  $\geq$  3 - Unacceptable  
# : Not Evaluated  
⌚ : Delayed Result Entry  
\* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416.  
Commercial Building - 1A,  
Kohinoor Mall, Kiroi Road, Kuria (W),  
Mumbai - 400070

End of Report

MINU CLINICAL LABORATORY (558)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 210410 and Sample No 02

Report Date : 02/11/2021

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable † : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_