



# Metropolis Healthcare Ltd., EQAS

MINU CLINICAL LABORATORY (558)

Outlier And Analyte Summary Report

Result Details For Cycle No 210207 and Sample No 01

Report Date : 19/07/2021

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	6,6.5
@ Specific Gravity	-	1.010	1.010,1.015,1.020
@ Glucose (urine)	-	2+	+1,+2,+3
@ Protein (Urine)	-	2+	+1,+2
@ Ketone (urine)	-	Negative	Negative
@ Blood (urine)	-	Negative	Negative
@ Bilirubin (urine)	-	Negative	Negative
@ Urobilinogen (urine)	-	Normal	Normal
X Nitrite (urine)	-	Absent	Present

Legend @ : Acceptable  
X : Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	1
EQAS Score Clinical Path - Urine Routine Chemistry	88.89 %

Dr Puneet Kumar Nigam  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416,  
Commercial Building - 1A,  
Door Mall, Kiroi Road, Kuria (W),  
Mumbai - 400070

End of Report

MINU CLINICAL LABORATORY (558)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 210207 and Sample No 01

Report Date : 19/07/2021

Analyte	Instrument	Result Value	Standard Unit
X Nitrite (urine)	Manual	Absent	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: Error - Analytical  
There is no sufficient evidence for Nitrite  
and Nitrite is not reported for UTR in Meru  
Clinical Lab

Corrective Action: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Reviewed by: 

Dated: 24/7/2021