

MINU CLINICAL LABORATORY (558)

Outlier And Analyte Summary Report

Result Details For Cycle No 210410 and Sample No 01

Report Date : 02/11/2021

Instrument : iCHROMA II

| Analyte | Standard Unit | Result Value | Accepted Value |
|---------------|---------------|--------------|----------------|
| X Dengue NS 1 | RU/ml | Positive | Negative |

Instrument : Manual

| Analyte | Standard Unit | Result Value | Accepted Value |
|---|---------------|------------------|------------------|
| @ HBsAg | IU/mL | Non-reactive | NR |
| @ HCV | IU/mL | Non-reactive | NR |
| # HIV Ab | S/CO | Non-reactive | |
| @ Typhidot IgM | - | Negative | Negative |
| @ Widal test - Typhi O antigen (TO) | - | No Agglutination | No Agglutination |
| @ Widal test - Typhi H antigen (TH) | - | No Agglutination | No Agglutination |
| @ Widal test - ParaTyphi A-H antigen (AH) | - | No Agglutination | No Agglutination |
| @ Widal test - ParaTyphi B-H antigen (BH) | - | No Agglutination | No Agglutination |

Instrument : Micro Lab 300

| Analyte | Standard Unit | Result Value | Accepted Value |
|---------------------|---------------|--------------|----------------|
| @ Rheumatoid Factor | IU/mL | Normal | Normal |

Instrument : Roche cobas c 111

| Analyte | Standard Unit | Result Value | Accepted Value |
|----------------------|---------------|--------------|----------------|
| @ C-Reactive Protein | mg/L | Abnormal | Abn |

Legend @ : Acceptable
X : Unacceptable
: Not Evaluated
⌚ : Delayed Result Entry

| | |
|------------------------------------|---------|
| Total Parameters | 11 |
| Not Evaluated Parameters | 1 |
| Evaluated Parameters | 10 |
| Outlier Parameters (X) | 1 |
| EQAS Score Qualitative Immunoassay | 90.00 % |

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416,
Commercial Building - 1A,
Kohinoor Mall, Kiroli Road, Kurla (W),
Mumbai - 400070

End of Report

| Analyte | Instrument | Result Value | Standard Unit |
|--|------------|--------------|---------------|
| X Dengue NS 1 | iCHROMA II | Positive | - |
| # HIV Ab | Manual | Non-reactive | Nil |
| Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry | | | |

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

MINU CLINICAL LABORATORY (558)
Outlier And Analyte Summary Report
Result Details For Cycle No 210410 and Sample No 02

Report Date : 02/11/2021

Instrument : ICHROMA II

| Analyte | Standard Unit | Result Value | Accepted Value |
|---------------|---------------|--------------|----------------|
| @ Dengue NS 1 | RU/ml | Negative | Negative |

Instrument : Manual

| Analyte | Standard Unit | Result Value | Accepted Value |
|---|---------------|------------------|------------------|
| @ HBsAg | IU/mL | Reactive | R |
| @ HCV | IU/mL | Non-reactive | NR |
| # HIV Ab | S/CO | Non-reactive | |
| @ Typhidot IgM | - | Negative | Negative |
| @ Widal test - Typhi O antigen (TO) | - | No Agglutination | No Agglutination |
| @ Widal test - Typhi H antigen (TH) | - | No Agglutination | No Agglutination |
| @ Widal test - ParaTyphi A-H antigen (AH) | - | No Agglutination | No Agglutination |
| @ Widal test - ParaTyphi B-H antigen (BH) | - | No Agglutination | No Agglutination |

Instrument : Micro Lab 300

| Analyte | Standard Unit | Result Value | Accepted Value |
|---------------------|---------------|--------------|----------------|
| @ Rheumatoid Factor | IU/mL | Normal | Normal |

Instrument : Roche cobas c 111

| Analyte | Standard Unit | Result Value | Accepted Value |
|----------------------|---------------|--------------|----------------|
| @ C-Reactive Protein | mg/L | Abnormal | Abn |

Legend @ : Acceptable
X : Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry

| | |
|------------------------------------|----------|
| Total Parameters | 11 |
| Not Evaluated Parameters | 1 |
| Evaluated Parameters | 10 |
| Outlier Parameters (X) | 0 |
| EQAS Score Qualitative Immunoassay | 100.00 % |

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416,
Commercial Building - 1A,
Kohinoor Mall, Kiroli Road, Kurla (W),
Mumbai - 400070

End of Report

**MINU CLINICAL LABORATORY (558)
 Outlier And Analyte Summary Report**

Outlier Details For Cycle No 210410 and Sample No 02

Report Date : 02/11/2021

| Analyte | Instrument | Result Value | Standard Unit |
|--|------------|--------------|---------------|
| # HIV Ab | Manual | Non-reactive | Nil |
| Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry | | | |

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____