

MINU CLINICAL LABORATORY (558)

Outlier And Analyte Summary Report

Result Details For Cycle No 210410 and Sample No 01

Report Date : 02/11/2021

Instrument : Roche cobas c 311

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ	
HbA1c	%	6.55	6.06	2.49	1.34	
<p>Legend @ : z score <math>\leq</math> 2.0 - Acceptable                      ! : 2.0 &lt; z score &lt; 3.0 - Warning                      X : z score <math>\geq</math> 3 - Unacceptable                      # : Not Evaluated                      ⌚ : Delayed Result Entry                      * : Not considered for evaluation.</p>						
					Total Parameters	1
					Not Evaluated Parameters	0
					Evaluated Parameters	1
					Outlier Parameters (X)	0
					EQAS Score HbA1C	100.00 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416,  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroli Road, Kurla (W),  
 Mumbai - 400070

End of Report

**MINU CLINICAL LABORATORY (558)**  
**Outlier And Analyte Summary Report**

**Outlier Details For Cycle No 210410 and Sample No 01**

**Report Date : 02/11/2021**

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! HbA1c	Roche cobas c 311	6.55	%	2.49
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

MINU CLINICAL LABORATORY (558)  
 Outlier And Analyte Summary Report  
 Result Details For Cycle No 210410 and Sample No 02

Report Date : 02/11/2021

Instrument : Roche cobas c 311

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ HbA1c	%	6.63	5.95	2.00	1.34

  

Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.	Total Parameters	1
	Not Evaluated Parameters	0
	Evaluated Parameters	1
	Outlier Parameters (X)	0
	EQAS Score HbA1C	100.00 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroi Road, Kurla (W),  
 Mumbai - 400070

End of Report

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable † : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_