

ARCRAY Hlth. Equip. Pvt. Ltd.
 Head Office
 20/20/1, 1st Floor
 Gokul, IPS Hill, Anantnagar
 District - 714 101
 Telephone: 091 2261111
 Bangalore: 08 446152996

Arkray Healthcare Pvt. Ltd. Installation Report:

To be filled by trained employee of Arkray Healthcare Pvt. Ltd.:

Instrument: PU Model No: 4010 Serial No: 62103008
 Date of Installation (DD/MM/YY): 04/08/2021 Time (HH:MM): 13:30
 Instrument under Warranty till: 04/08/2024 (03 year from date of installation)
 Installed by: Rajesh Kumar Yadav
 Installation Details (mention OK if everything is alright or else the status observed):
 Box Content: OK
 Power Supply: OK
 Instrument Display: OK
 Button Panel: OK
 Printer: OK
 Demo: Done

To be filled by Customer:

Name: Minu Clinical Laboratory
 Address: Shop no 7/8 Sunflower bldg, Royal complex, Panch Road, Bosiswadi
 Phone: 9820181968 Email: airaglab.minulab@gmail.com

Satisfaction status about the Installation: Satisfied / Dissatisfied (Reason) _____


Satisfaction status about the demo: Satisfied / Dissatisfied (Reason) _____

For Arkray Healthcare Pvt. Ltd. For (Customer) _____

Signature: [Signature]
 Name: Rajesh Kumar Yadav
 Phone: 9892783412

Signature: [Signature]
 Name & Stamp: Dr. Anirag B. Sach
 Phone: 9820181968





arkray healthcare Pvt. Ltd.

Plot no. 336, 338, 340, Road no. 3, G.I.D.C.,
Sachin 394 230, (Surat), India. Telephone: 0261-239 7712 to 15
Facsimile: 0261-239 7719, URL <http://www.arkray.co.in>

Installation / Operation/ Performance Qualification Certificate.

Pocket Chem UA (Urine Analyzer)

System Description:

Type of Analyzer	Pocket Chem UA (Urine analyzer)
Model	PU-4010
Serial Number	62103008
Manufactured By	ARKRAY INC, JAPAN
Marketed in India By	ARKRAY HEALTHCARE PVT LTD.

Instrument Delivery:

Date of Delivery	04/08/2021
-------------------------	------------





arkray healthcare Pvt. Ltd.

Plot no. 336, 338, 340, Road no. 3, G.I.D.C.,
Sachin 394 230, (Surat), India. Telephone: 0261-239 7712 to 15
Facsimile: 0261-239 7719, URL <http://www.arkray.co.in>

Installation Qualification Certificate.


Pocket Chem PU-4010	Received	Undamaged
Accessories	Received	Undamaged

Assembly and Installation:

Place of Installation	Minu Clinical Laboratory (Mumbai)
Date of Installation : 04-08-2021 Analyzer Serial No : 62103008 Installed By : Mr.Rajesh Kumar Yadav	
Initial Check Up / Auto Self Diagnosis	Passed
Pocket Chem PU-4010 successfully installed and is in a fully working condition.	
  For, Arkray Healthcare Pvt Ltd.	Customer's Signature and Stamp

Regd. Office: 701/702, Opulence.6th Road, TPS III, Santacruz(East) Mumbai-400 055 INDIA :telephone: +91 - 22-6155 9355 : facsimile: +91-22-61559356:

CIN NO :U51507MH2007PTC175112



arkray healthcare Pvt. Ltd.

Plot no. 336, 338, 340, Road no. 3, G.I.D.C.,
Sachin 394 230, (Surat), India. Telephone: 0261-239 7712 to 15
Facsimile: 0261-239 7719, URL <http://www.arkray.co.in>



PERFORMANCE QUALIFICATION (PO)

Analyzer: Pocket Chem UA (PU-4010)
Serial No: 62103008

Date : 27-09-2021

Verification of result is done by running Check strip.

Check Strips run result									
GLU	PRO	BIL	URO	PH	SG	BLD	KET	NIT	LEU
(-)	(-)	(-)	(-) Normal	5.0	>1.030	0.06 (+)	(-)	(-)	(-)

Check strip run on the PU4010 and results are found to be within range	
 For, Arkray Healthcare Pvt Ltd.	 Customer's Signature and Stamp

Regd. Office: 701/702, Opulence.6th Road, TPS III, Santacruz(East) Mumbai-400 055 INDIA : telephone: +91 - 22-6155 9355 : facsimile: +91-22-61559356:

CIN NO : U51507MH2007PTC175112



arkray healthcare Pvt. Ltd.

Plot no. 336, 338, 340, Road no. 3, G.I.D.C.,
Sachin 394 230, (Surat), India. Telephone: 0261-239 7712 to 15
Fascimile: 0261-239 7719, URL <http://www.arkray.co.in>


OPERATIONAL QUALIFICATION (OQ)

Analyzer: Pocket Chem UA (PU-4010)
Serial No: 62103008

Date : 27-09-2021

OPERATIONAL PARTS	Working	Not Working
Display on Switch On	O.K	---
Keypad Panel	O.K.	---
Strip transportation Movement	O.K.	---
Self-test on Power ON	O.K.	---
Printer	O.K.	---

All the parts of Pocket Chem PU 4010 are in a working condition & Operation of analyzer is according to given specifications at the user's environment.

 <p>For, Arkray Healthcare Pvt Ltd.</p>	<p>Customer's Signature and Stamp</p>
---	--

Regd. Office: 701/702, Opulence.6th Road, TPS III , Santacruz(East) Mumbai-400 055 INDIA :telephone: +91 – 22-6155 9355 : facsimile: +91-22-61559356:

CIN NO : U51507MH2007PTC175112