

**CALIBRATION CERTIFICATE**

Ref. No: BI/CAL/21/4972

Date of Issue: 15.03.2021

Name of customer: NATIONAL AYURVEDHA RESEARCH INSTITUTE FOR

PANCHAKARAMA, CHERUTHURUTHY

<b>Equipment</b>	<b>Make/Model</b>	<b>Serial No</b>
LABORATORY CENTRIFUGE	REMI /R - 8C PLUS	ZHBN-07595
<b>Date of Calibration</b>	<b>Calibration Due On</b>	
15.03.2021	14.03.2022	

*Certified that the above instrument has been calibrated by trained technical personnel of this organization using measuring equipment traceable to National Standards.*

**TEST AND CALIBRATION DATA****Parameter : RPM**

Room temperature at the time of Calibration: 22°C

Range	Standard Reading	DUC* Reading	Error	Remarks
6000	1500	1497.0	3.0	OK
	2500	2498.0	2.0	OK
	3500	3497.0	3.0	OK
	4500	4946.0	4.0	OK


\*DUC- Device Under Calibration

Acceptable error as per: DC Act 1940/ Manufacturer Manual:

Temperature	Weight	Speed
+/- 2 Degree C	+/- 2% gm	+/- 20 RPM

Details of Master Equipment used:

Description	Make/Model	Serial No	Traceable to
Digital Non-Contact Tachometer	FISHERBRAND	192096582	National Standards
Calibrated at	Certification No	Calibration Due Date	
STIC, EKM	MM/210811B/01	18-08-2022	

Prepared By: Checked By: 

Door No.5/302 A, Nechoor Road, Nechoor P.O, Ernakulam – 686664

Tel : 0484 – 2316938, Email : info@inlabinstruments.com Web :

www.inlabinstruments.com