


LAB MONTHLY SUMMARY

Lab Name **DR DAS PATH LAB**Lab No **9006**Month **November**Year **2021**Constituent Group **Chemistry I**Date of Result Entered : **15/11/2021**Date of Report Published : **03/12/2021**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Transasia / Erba	846	148.72	6.42	9.54	143.4 mg/dl	-0.56	0.66
2	UREA	Urease UV / GLDH	Transasia / Erba	775	171.11	8.24	14.11	177.7 mg/dl	0.47	1.01
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Transasia / Erba	425	3.58	9.55	0.34	3.53 mg/dl	-0.15	0.03
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Transasia / Erba	803	3.81	10.30	0.39	3.03 mg/dl	-1.99	0.03
5	T-PROTEIN	Biuret - Colorimetric	Transasia / Erba	811	5.12	7.68	0.39	5.06 g/dl	-0.15	0.03
6	ALBUMIN	BCG - colorimetric	Transasia / Erba	826	2.99	5.55	0.17	2.92 g/dl	-0.42	0.01
7	CALCIUM	Arsenazo III	Transasia / Erba	580	9.75	8.68	0.85	8.7 mg/dl	-1.24	0.07
8	URIC ACID	Enzymatic / Uricase Colorimetric	Transasia / Erba	789	3.93	11.57	0.46	3.5 mg/dl	-0.95	0.03
9	CHOLESTEROL	CHOD-PAP	Transasia / Erba	821	93.58	9.03	8.45	90 mg/dl	-0.42	0.59
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Transasia / Erba	806	113.91	10.20	11.62	99.6 mg/dl	-1.23	0.82
11	HDL CHO	Direct method / Enzymatic colorimetric	Transasia / Erba	617	24.25	11.78	2.86	21.8 mg/dl	-0.86	0.23
12	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	620	126.99	3.58	4.55	119 mmol/L	-1.76	0.37
13	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	622	6.05	7.80	0.47	6.2 mmol/L	0.32	0.04
14	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	485	93.12	5.76	5.37	83 mmol/L	-1.89	0.49
15	AST	UV kinetic without PLP (P-5-P)	Transasia / Erba	591	59.67	10.21	6.09	56.6 U/L	-0.50	0.50
16	ALT	UV kinetic without PLP (P-5-P)	Transasia / Erba	569	121.06	12.64	15.30	80.1 U/L	-2.68	1.28
17	ALP	PNP AMP kinetic	Transasia / Erba	669	263.78	15.38	40.56	221 U/L	-1.05	3.14

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.

Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

******* End of Report *******