

CERTIFICATE OF CALIBRATION VERIFICATION

Customer Name : Pathology Lab, District Women Hospital, Pilibhit (U.P.)

Instrument Name : Selectra (Fully Automated Biochemistry Analyzer)

Model : Pro M

Serial Number : 16-4152

Calibration done on : 24-12-2021

Next Calibration due : 23-12-2022

Contact Person: Dr. Afzal Husain

This is to certify that during its scheduled PMS above mentioned instrument has been verified for instrument calibration. During instrument calibration Lamp check, filter check, Pipetter unit dispense check and potassium dichromate tests were performed and found complying for all the parameters as per the specifications provided by the manufacturer. After instrument mechanical calibration verification, individual tests were calibrated and controls were run for which results obtained were within the specified limits.



Calibration and PMS done by : Gajendra Pratap Singh
Service Engineer
POCT Services