

**CERTIFICATE OF CALIBRATION VERIFICATION**

**Customer Name :** Pathology Lab, District Women Hospital, Pilibhit (U.P.)

**Instrument Name :** Medonic M series      **Model :** M20      **Serial Number :** 1400062

**Calibration done on :** 24-12-2021      **Next Calibration due :** 23-12-2022

**Contact Person:** Dr. Afzal Husain

This is to certify that during its scheduled PMS above mentioned instrument has been verified for its calibration for six CBC parameters i.e. RBC, MCV, PLT, MPV, HGB and WBC in open tube mode after verifying the background run for the machine. Calibration was verified by running three levels of control.



**Calibration and PMS done by :** Gajendra Pratap Singh  
Service Engineer  
POCT Services