

Enabling Better Outcomes

CALIBRATION CERTIFICATE

This is to certify that Biochemistry Analyser **SF-3000 PLUS** bearing **Serial No.4512-1493** which has been installed at **ANNAI DIAGNOSTIC LAB, KULLANCHAVADI** has been tested on the following aspects

1. LIGHT SOURCE :

Constant Light intensity for each filter.

2. CALIBRATION :

The Absorbance Factor according to the Factory set calibration.
The Read Cell Temperature according to the Factory set calibration.

3. FILTERS :

FILTERS(nm)	VOLTS(v)
340	3.90
405	5.00
505	5.90
545	6.30
580	6.30
630	6.50

Quality of the individual filters viz.340, 405, 505, 545, 580 and 630 were checked for calibration through the instrument software and calibrated.

On complete testing of the unit we found that the unit is working satisfactorily and used to carry out the necessary functions.

Date of Calibration : 25-08-2021
Next Calibration Due : 24-08-2022

For **CPC DIAGNOSTICS PVT LTD**

Authorised Signatory



CORPORATE OFFICE

No.70/6, Old No.108/6, 4th floor, Westminster, Dr. Radhakrishnan Salai, Mylapore, Chennai - 600 004.
Phone No: 044-23460168/169