



Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area
Delhi - 110092

Phone No. : 09810943334 011-41764006, Email: multitech.technofab@gmail.com, mtedelhi2020@gmail.com,
Web. : www.multitechtechnofab.com



CC3225

CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/1612 Certificate Issued on: 03-08-2021

Job / Identification No. : MTE/CAL/JOB/2460 Date of Calibration : 02-08-2021

Job Request Date: 30-07-2021 Recommended Date of Next Calibration: 02-08-2022

Calibration Certificate for : Centrifuge Item Location : NA

CALIBRATED FOR YASH DIAGNOSTIC CENTRE
702,VISHESHARPUR,PACHHATIA, Azamgarh Rd, , JAUNPUR, Uttar Pradesh,
222001 Reference:
Order thru Telephonic

DISCIPLINE Mechanical GROUP Acceleration & Speed

Item's Description

Name of the Item	Make / Model	Serial / I.D. No.	Working / Inst. Range	Unit	Item Condition
Centrifuge	REMI / NA	NA / YASH/CF/1	0 to 4000	RPM	Working

ENVIRONMENTAL CONDITION	Temperature (in °C):	25.8	± 2 °C	Humidity (in % RH):	51	± 10% RH
-------------------------	----------------------	------	--------	---------------------	----	----------

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make / Model	Certificate No.	Calibration Agency / Traceability	Date of Calibration	Valid up to
Digital Tachometer(Non contact)	Lutron / DT-2234C	2101/19/L/001	QTEX Consumer Products, Fbd	18-01-2021	18-01-2022

PRINCIPLE/METHODOLOGY OF CALIBRATION: Lab Procedure No.: MTE-2/RPM/2.50 St. Procedure No.: NA

RESULTS: On-Site Calibration
All Readings are in RPM, Specified Otherwise

U.U.C. READING(Set At)	Average	UNCERTAINTY^ in (%)
Knob1	957.7	± 0.48
Knob2	1422	± 0.74
Knob3	2228	± 0.48
Knob4	2931.3	± 0.48
Knob5	3875.3	± 0.48

^The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.

REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.

Note

- This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.
- The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor implied.
- Total liability of our organization is limited to the invoiced amount.
- Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
- In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

***** END OF REPORT *****

For and on behalf of
Multitech Technofab Enterprises

ULR : CC322521000001063F

Calibrate By
(Parveen Kumar) , parveen
Deputy Technical Manager


Laboratory Head