



Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area
Delhi - 110092

Phone No. : 09810943334 011-41764006, Email: multitech.technofab@gmail.com, mtedelhi2020@gmail.com, Web. :
www.multitechtechnofab.com



CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/3741			Certificate Issued on: 07-01-2022		
Job / Identification No. : MTE/CAL/JOB/4931		Date of Calibration : 07-01-2022			
Item Receipt Date: 06-01-2022		Recommended Date of Next Calibration: 07-01-2023			
Calibration Certificate for : Digital Temperature Indicator with sensor					Item Location : Lab
Calibrated For	YASH DIAGNOSTIC CENTRE 702,VISHESHARPUR,PACHHATIA, Azamgarh Rd,			Reference: Order thru Telephonic	
Discipline	Thermal	Group	Temperature		
Item's Description					
Name of the Item	Make / Model	Serial / I.D. No.	Range	Item Condition	
Digital Temperature Indicator with sensor	SYMBOL / NA	NA / NA	Mentioned hereunder	Working	
Environmental Condition	Temp (in °C):	26.5	± 0.5 °C	RH (in % RH):	26.7 ± 3% RH
STANDARD REFERENCE DETAILS Traceable to National / International Standard					
Name of the Standard	Make / Model	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
4 Wire RTD (PT-100) Sensor with Indicator	Tempens and Lutron / PT-100 and TM-917	ULR-CC288821000001903F	Sigma Test & Research Centre	18-12-2021	18-12-2022
USED EQUIPMENT DETAILS		1. Methanol Bath			
Calibration Procedure	Lab Procedure : MTE-2/T/2.4		STD Procedure : DKD-R-5-1 & EURAMET CG-8		

RESULTS: In-Lab Calibration
All Readings are in °C, Specified Otherwise

SET AT	U.U.C. READING (AVERAGE)	CORRECTED VALUE OF STANDARD (AVERAGE)	UNCERTAINTY [^]
Temperature	Range -50 to 80 °C Least Count 0.1		
-20 °C	-20.0 °C	-20.03 °C	± 0.24 °C
-10 °C	-10.0 °C	-10.05 °C	± 0.24 °C
2 °C	2.0 °C	2.05 °C	± 0.24 °C
4 °C	4.0 °C	4.02 °C	± 0.24 °C
8 °C	8.0 °C	8.01 °C	± 0.24 °C

[^]The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.

REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.

Note

- This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.

For and on behalf of
Multitech Technofab Enterprises

ULR : CC322522000000132F

Calibrate By Sachin
Deputy Quality Manager


Laboratory Head



CC3225

Multitech Technofab Enterprises475, First Floor, F.I.E., Patparganj Industrial Area
Delhi - 110092Phone No. : 09810943334 011-41764006, Email: multitech.technofab@gmail.com, mtedelhi2020@gmail.com, Web. :
www.multitechtechnofab.com**CALIBRATION CERTIFICATE**

Calibration Certificate No.: MTE/204/CC/3741

Certificate Issued on: 07-01-2022

2. The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor implied.
3. Total liability of our organization is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

***** END OF REPORT *****

ULR : CC32252200000132F

Calibrate By Sachin
Deputy Quality ManagerFor and on behalf of
Multitech Technofab Enterprises

Laboratory Head