

LAB MONTHLY SUMMARY



Lab Name **RAO'S STANDARD DIAGNOSTIC PVT LTD** Lab No **13937**
 Month **December** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 21/12/2021

Date of Report Published : 03/01/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	969	215.80	9.25	19.96	215 mg/dl	-0.04	1.28
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	533	107.40	11.34	12.17	73.06 mg/dl	-2.82	1.05
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	725	2.24	11.51	0.26	2.49 mg/dl	0.97	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	789	3.00	13.68	0.41	2.5 mg/dl	-1.22	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	738	5.21	9.17	0.48	5.3 g/dl	0.19	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	728	3.19	10.00	0.32	3.95 g/dl	2.38	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	776	4.39	19.07	0.84	4.65 mg/dl	0.31	0.06
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	975	108.50	11.43	12.40	114 mg/dl	0.44	0.79
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	803	96.54	15.20	14.68	99.7 mg/dl	0.22	1.04

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****

LAB MONTHLY SUMMARY



Lab Name **RAO'S STANDARD DIAGNOSTIC PVT LTD** Lab No **13937**
 Month **November** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 20/11/2021

Date of Report Published : 03/12/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1004	144.70	10.23	14.80	139 mg/dl	-0.39	0.93
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	577	42.79	12.32	5.27	40.01 mg/dl	-0.53	0.44
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	709	9.02	12.78	1.15	10.44 mg/dl	1.23	0.09
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	940	2.43	23.94	0.58	2.14 mg/dl	-0.50	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	768	5.44	10.95	0.60	5.25 g/dl	-0.32	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	760	3.22	9.92	0.32	3.75 g/dl	1.66	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	801	8.10	14.01	1.14	8.15 mg/dl	0.04	0.08
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	993	112.61	12.13	13.66	104 mg/dl	-0.63	0.87
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	843	156.72	12.90	20.21	165 mg/dl	0.41	1.39

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:


Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****

LAB MONTHLY SUMMARY



Lab Name **RAO'S STANDARD
DIAGNOSTIC PVT LTD** Lab No **13937**
Month **October** Year **2021**
Constituent Group **Chemistry II**



Date of Result Entered : 20/10/2021

Date of Report Published : 01/11/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1038	107.73	10.13	10.92	99 mg/dl	-0.80	0.68
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	612	28.11	15.98	4.49	24.09 mg/dl	-0.90	0.36
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	744	1.32	12.58	0.17	1.49 mg/dl	1.02	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	801	7.05	13.48	0.95	7.14 mg/dl	0.09	0.07
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	793	5.35	10.66	0.57	5.81 g/dl	0.81	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	769	3.19	10.00	0.32	3.22 g/dl	0.09	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	833	5.90	15.69	0.93	7.03 mg/dl	1.22	0.06
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	908	106.08	11.36	12.05	146.14 mg/dl	3.32	0.80
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	754	101.65	15.55	15.81	90 mg/dl	-0.74	1.15

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Co-Ordinator
Christian Medical College, Vellore

***** End of Report *****

