

## LAB MONTHLY SUMMARY



Lab Name ANUSAKTHI MICROLAB Lab No 13883  
 Month July Year 2021  
 Constituent Group Chemistry II



Date of Result Entered : 19/07/2021

Date of Report Published : 02/08/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	986	147.82	9.98	14.75	154.4 mg/dl	0.45	0.94
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation)	853	34.55	15.04	5.20	37.2 mg/dl	0.51	0.36
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	720	2.75	11.18	0.31	3.2 mg/dl	1.46	0.02
4	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	742	5.53	9.90	0.55	6.1 g/dl	1.04	0.04
5	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	711	3.31	9.70	0.32	3.5 g/dl	0.59	0.02
6	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	834	5.17	19.21	0.99	5.3 mg/dl	0.13	0.07
7	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	984	115.76	13.00	15.04	112.7 mg/dl	-0.20	0.96
8	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	818	177.61	12.70	22.56	180.4 mg/dl	0.12	1.58

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*