

VIEW LAB MONTHLY SUMMARY

Lab Name	ANUSAKTHI MICROLAB	Details About Robust Analysis	
Lab No	13883	Detail About Monthly Summary	
Month	August	Detail about SDI	
Year	2021		
Constituent Group	Chemistry II		
Click on the analyte to view Graphical Data			
		All Analyser Result	Print
Date of Result Entered : 19/08/2021			
Date of Report Published : 01/09/2021			

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	918	247.66	10.88	26.94	258.1 mg/dl	0.39	1.78
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	873	44.63	13.73	6.13	50.6 mg/dl	0.97	0.41
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	702	1.90	13.28	0.25	2 mg/dl	0.4	0.02
4	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	727	5.45	10.82	0.59	5.4 g/dl	-0.08	0.04
5	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	672	3.29	10.23	0.34	3.4 g/dl	0.32	0.03
6	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	764	3.93	24.71	0.97	4.3 mg/dl	0.38	0.07
7	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	895	112.13	11.42	12.81	105.5 mg/dl	-0.52	0.86
8	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	784	130.89	14.50	18.98	143.7 mg/dl	0.67	1.36

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)[Yearly Summary](#)[Change Password](#)[Home](#)[Sign-out](#)