

LAB MONTHLY SUMMARY



Lab Name ANUSAKTHI MICROLAB Lab No 13883
 Month October Year 2021
 Constituent Group Chemistry II



Date of Result Entered : 20/10/2021

Date of Report Published : 01/11/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1038	107.73	10.13	10.92	117.4 mg/dl	0.89	0.68
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	887	29.20	15.70	4.58	31.6 mg/dl	0.52	0.31
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	744	1.32	12.58	0.17	1.5 mg/dl	1.08	0.01
4	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	793	5.35	10.66	0.57	6 g/dl	1.14	0.04
5	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	769	3.19	10.00	0.32	3.5 g/dl	0.97	0.02
6	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	833	5.90	15.69	0.93	5.7 mg/dl	-0.22	0.06
7	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	908	106.08	11.36	12.05	91.6 mg/dl	-1.20	0.80
8	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	754	101.65	15.55	15.81	119.4 mg/dl	1.12	1.15

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****