

LAB MONTHLY SUMMARY



Lab Name ANUSAKTHI MICROLAB Lab No 13883
 Month November Year 2021
 Constituent Group Chemistry II



Date of Result Entered : 20/11/2021

Date of Report Published : 03/12/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1004	144.70	10.23	14.80	159.5 mg/dl	1.00	0.93
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	844	43.73	12.85	5.62	44.6 mg/dl	0.15	0.39
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	709	9.02	12.78	1.15	9.9 mg/dl	0.76	0.09
4	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	768	5.44	10.95	0.60	5.4 g/dl	-0.07	0.04
5	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	760	3.22	9.92	0.32	3.3 g/dl	0.25	0.02
6	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	801	8.10	14.01	1.14	8.8 mg/dl	0.62	0.08
7	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	993	112.61	12.13	13.66	125.4 mg/dl	0.94	0.87
8	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	843	156.72	12.90	20.21	166.1 mg/dl	0.46	1.39

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****