

LAB MONTHLY SUMMARY



Lab Name **ANUSAKTHI MICROLAB** Lab No **13883**
 Month **December** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 20/12/2021

Date of Report Published : 03/01/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	969	215.80	9.25	19.96	225.4 mg/dl	0.48	1.28
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	762	105.93	13.00	13.77	117.8 mg/dl	0.86	1.00
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	725	2.24	11.51	0.26	2.4 mg/dl	0.62	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	789	3.00	13.68	0.41	3.6 mg/dl	1.46	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	738	5.21	9.17	0.48	5.4 g/dl	0.40	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	728	3.19	10.00	0.32	3.6 g/dl	1.29	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	776	4.39	19.07	0.84	5 mg/dl	0.73	0.06
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	975	108.50	11.43	12.40	112.4 mg/dl	0.31	0.79
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	803	96.54	15.20	14.68	103.4 mg/dl	0.47	1.04

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****