




# ANNUAL MAINTENANCE CONTRACT REPORT

CUSTOMER NAME: <u>ANUSAKTHI MICRO LAB</u>		DATE	<u>29</u>	<u>12</u>	<u>2021</u>
Address: <u>No, 5/118N, KMS TUS OPP, SALEM S2 BUS STOP NEAR, LAXMI NAGAR</u>					
District: <u>ERODE</u>	State: <u>TN</u>				
Pin: <u>638 316</u>	Ph. No: <u>89731 88990</u>	Email ID:			
Instrument: <u>AUTO LAB</u>	Model: <u>VERBA</u>				
Serial No: <u>ES1021705272</u>					
Instrument Installation Date:					
Contact Person:					
S/N	NAME	DESIGNATION	PHONE/MOBILE NO.	EMAIL ID	
	<u>N. RAJ KUMAR</u>	<u>OWNER</u>	<u>8973188990</u>	<u>anusakthimicrolab@gmail.com</u>	
Contract steps to be followed as per Company Norms:					
<ul style="list-style-type: none"> <li>- This Annual Maintenance Period is valid for one year from date of Maintenance contract to be signed.</li> <li>- Annual Maintenance Charge excludes the price of spares/consumables if required. It is only for servicing / preventive maintenance / application support/courtesy visit.</li> <li>- During AMC period our engineer will visit and check the instrument status for 4 times in a year and will inform you the necessary steps to be taken. This may be a preventive maintenance/ service support/ application support.</li> </ul>					
AMC / C-AMC Details:					
AMC / C-AMC Period (1 year) From: <u>29-12-2021</u> To: <u>28-12-2022</u>					
AMC / C-AMC Amount Collected: Rs. <u>10000+437 = 11,800</u> , AMC Amount Collected (In words): <u>Eleven thousand Eight hundred only.</u>					
Payment Details (DD/Cheque No.):					
Current Status of the Instrument:					
NOTE: Enclose the current dated Technical Service Report as attachment.					
ENGINEER'S REMARK: <u>Machine working good</u>					
SERVICE ENGINEER NAME:			Seal & Signature of the Customer		
SIGNATURE: <u>S. Latha Sowtham</u>					
DATE OF VISIT (DD/MM/YY): <u>29.12.2021</u>			Name: <u>ANUSAKTHI MICRO LAB</u>		
			(with Seal) <u>5/118N, K.M.S. Complex</u>		
			Date: <u>SALEM BUS STOP, LAKSHMI NAGAR,</u>		
			<u>VASAVI COLLEGE-638316</u>		
			<u>ERODE DT.</u>		
NOTE: Please fill the form in Capital Letters only.					
IMP: Please note that in case of non-compliance of above points, report shall not be accepted at H.O.					

**Peerless Biotech Pvt Ltd.**

36/37 'Millenium House', M.K. Srinivasan Nagar Main Road, Perungudi, Chennai-600 096, Tamil Nadu, INDIA  
Mobile : 8939991003 Ph : 044-2496 0541, Email : instruments@peerlessbio.com, Website : www.peerlessbio.com



## CERTIFICATE OF CALIBRATION

Calibration Date : 29-12-2021

Next Calibration: 28-12-2022

Certificate No : 202111184

Name of the Instrument : Chemistry Analyser

Model : Auto Lab Versa

Make : E-LAB Biological Science & Technology Co., Ltd.

Serial No : ES1021705272

Customer Address : Anusakthi Micro Lab  
No.5/118 N.K.M.S Complex, Salem Bus Stop,  
Lakshmi Nagar, Vasavi College - 638316. Erode Dt.

Calibrated by:

Sounderarajan



Reviewed by:

Karthikeyan

NOTE: Only for Audit Purpose.

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. Anusakthi Micro Lab, 5/118 N, KMS Complex, Salem Service Road, Bus Stop, Lakshmi Nagar, Erode - 638 316.	Certificate No:	AMS/LAB/2889696
		Issuing Date:	14/08/2021
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Bio Chemistry Analyser	Id. No:	ASM/LAB /BCA/001
Make:	VERSA	Department:	Laboratory
Model:	Auto Lab	Calibrated On:	13/08/2021
Sl. No:	ES1021705272	Calibration Due Suggested:	12/08/2022
Type:	Semi Auto	DUC Condition on Receipt:	OK
Range & Least Count:	---		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	RigelSafeTest60	PSG/CIRD/BMCC/5507/2021	18/01/2022
2	Thermo Hygrometer	Beetech/088	PC/01/21/ATL/0194-01	22/01/2022

Room Temperature: 25°C±2 Room Humidity: 30-60%

(At the time of Calibration)

### ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	0.368	mA	500	Normal
2	Frequency	50.0	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	232.2	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	233.3	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	4.2	Volts	< 7	Normal

\*Electrical Safety: Please check the attached log sheet

SL.No	Parameter	Measured	Standard	Correction	Unit
1	Lamp Voltage	11.68	12.00	-0.32	Volts
2	Detector Sensitivity	-0.00	0.0	0.0	% Absorbance
3	Sipper Volume	404.0	450	-46.0	µl
4	Peristaltic Pump pressure	-13.47	-15	+1.53	cmH <sub>2</sub> O

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

B. S. S.



Approved by:

(Technical Manager)

B. S. S.